



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

WAB

Agency: Miami Dade P.D.

Instrument Serial Number: 80-00886

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
<input type="checkbox"/> Registration <input type="checkbox"/> Request for Registration Attached <input type="checkbox"/> Annual <input type="checkbox"/> Inspection After Repair <input type="checkbox"/> Other: _____ <input type="checkbox"/> Barometric Pressure: _____ Gauge _____ Instrument <input type="checkbox"/> Instrument Set Up Verified	<input checked="" type="checkbox"/> Agency Inspection Notification <input checked="" type="checkbox"/> Agency Contact <input type="checkbox"/> Other: _____ Date of Notification/Contact: 2/14/13 Agency Inspector: Pablo Espinosa	The following records were audited: <input type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: _____ <input checked="" type="checkbox"/> Other: _____

Comments:  
 Pablo Espinosa notified me of an issue with 880 not reading the first acetone reading. After troubleshooting the issue it was determined that he would be sending the instrument in for repair to Enforcement Electronics.

Instrument/Area:	Equipment:	Supplies:
<input type="checkbox"/> Clean/Dry <input type="checkbox"/> Secure <input type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	<input type="checkbox"/> Distilled/Deionized Water <input type="checkbox"/> Mouth Alcohol Solution <input type="checkbox"/> Acetone Stock Solution <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard <input type="checkbox"/> Mouth Pieces

Comments:

Action:	Comments:
<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Remain in Evidentiary Use <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remove from Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	    

Signature of Alcohol Testing Program Staff Member

2/15/13 Date

Original - FDLE Copy - Agency