Florida Department of Law Enforcement
Alcohol Testing Program PGM

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE
Time of Inspection: 10:25 Date of Inspection: 03/30/2015

Serial Number: 80-000224
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:
SKIPPED AI TO OPERATE INSTRUMENT

N/A AI not conducted
compliance not determined

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JENNIFER H JACOBS
Signature and Printed Name

03/30/2015
Date

FDLE/ATP Form 40 – March 2004