

RECEIVED

DEC 28 2015

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - DRICKLYZER 8000

Agency: POLK COUNTY SO
Inspector: [Name]

Date of Inspection: 12/21/2015

Serial Number: 80-0017
Inventory: 0100.27

Check of Test	98%	98%
Date Recal/Rec Time Adjusted		
Pre-Inspection (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Alcohol Control Test: 0.100		
Interferent Control Test: Interferent Detect		
Post-Inspection (Post-Inspection): OK		

Alcohol Free Subject Test	0.05g/210L Test (g/210L) Lot#: 201403E Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201403E Exp: 03/25/2016	0.10g/210L Test (g/210L) Lot#: 201403E Exp: 03/25/2016	0.05g/210L Test (g/210L) Lot#: 201403E Exp: 03/25/2016
0.000	0.048	0.079	0.115	0.078
0.000	0.040	0.079	0.107	0.078
0.000	0.045	0.079	0.107	0.078

Inspector Signature: [Signature]
Inspector Name: [Name]

This device is a breathalyzer () does not comply () with Chapter 310.04, F.S.
The user shall hold a valid Florida Department of Law Enforcement Agency Inspector License and shall maintain the device in accordance with the provisions of Chapter 310.04, F.S.

James A. Pearson
Signature and Printed Name

12/21/2015
Date

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

NOV 02 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DULY COUNTY 20

Time of Inspection: 14:53

Date of Inspection: 10/29/2015

Serial Number: 86-000795

Software: 8100.27

Check for Yes	Yes	NO
Is the subject's eye adjusted		
Highway Check (Pre-Inspection): OK		OK
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Y/N	
Interferent Detect Test: Interferent Detect	Yes	
Diagnose to Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/09/2016	0.20g/210L Test (g/210L) Lot#: 201406G Exp: 07/11/2016	0.05 g/210L Dry Gas Seal Test (g/210L) Lot#: 145150807 Exp: 09/03/2017
0.000	0.051	0.079	0.198	0.078
0.000	0.051	0.079	0.198	0.078
0.000	0.051	0.079	0.199	0.078

Number of breaths tested: 1

Remarks:

The law enforcement complies () does not comply (), with Chapter 316.04.

I, the undersigned, hold a valid Florida Department of Law Enforcement Agency Inspector's License and am performing this report in accordance with the provisions of Chapter 316.04, F.S.

James A. Newman

Signature and Printed Name

10/29/2015
Date

RECEIVED

SEP 28 2015

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 15:25

Date of Inspection: 09/21/2015

Serial Number: 80-000795
Software: 8100.57

Check or Test	Yes	No
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201402F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201401C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201403E Exp: 06/12/2016	0.05 g/210L Dry Gas Std Test (g/210L) Lot#: 15A.50000 Exp: 09/30/2017
0.000	0.052	0.080	0.196	0.078
0.000	0.052	0.081	0.198	0.078
0.000	0.052	0.081	0.198	0.073

Remarks: (Optional) _____

Remarks:

The above instrument complies () does not comply () with Chapter 119-8, F.S.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 119-8, F.S.

James A. Newman

JAMES A. NEWMAN
Signature and Printed Name

09/21/2015
Date

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

SEP 03 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BUREAU COUNTY 9
Time of inspection: 10:16

Date of Inspection: 08/21/2015

Serial Number: 80-000795
Software: 8100.27

Check for Pass	YES	NO
Visual and/or Physical Inspection		No
Alcohol Check (Pre-Inspection): OK		
Alcohol Free Sample Test: 0.000	Yes	
Alcohol Free Sample Test: Slope Not Met	Yes	
Alcohol Free Sample Test: Interferent Detect	Yes	
Alcohol Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201404C Exp: 06/08/2016	0.10g/210L Test (g/210L) Lot#: 201404C Exp: 06/08/2016	0.20g/210L Test (g/210L) Lot#: 201404C Exp: 06/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 29814080A3 Exp: 11/05/2015
0.000	0.080	0.081	0.187	0.079
0.000	0.080	0.081	0.189	0.079
0.000	0.080	0.081	0.200	0.079

Number of Significant Figures: 1

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newman

Signature and Printed Name

08/21/2015
Date

RECEIVED

AUG 05 2015

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

INSPECTION REPORT - JETSETZER 8000

Agency: FLORIDA DEPT OF LAW ENFORCEMENT
Inspector: DAVID L. WILSON

Date of Inspection: 07/31/2015

Serial Number: 80-050795
Software: 8109.27

Calibrated	YES	NO
Alcohol Test: Adjusted		
Alcohol Test: Sample Error: 0.000	YES	NO
Alcohol Test: Slope Not Met	YES	NO
Interference: Detect	YES	NO
Interference: Detect	YES	NO
Signature: (Post-Inspection): OK	YES	NO

Alcohol Test (g/210L)	0.05g/210L Test (g/210L)	0.08g/210L Test (g/210L)	0.10g/210L Test (g/210L)	0.15g/210L Test (g/210L)
0.000	0.050	0.082	0.100	0.150
0.000	0.050	0.082	0.100	0.150
0.000	0.050	0.082	0.100	0.150

Inspector: DAVID L. WILSON

Signature:

This instrument complies () does not comply () with Department of Law Enforcement.

This instrument is a valid Florida Department of Law Enforcement model. This instrument is in accordance with the provisions of Chapter 318.02, F.S.

James A. Pearson
Signature and Printed Name

Signature and Printed Name

07/31/2015
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUL 07 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILL COUNTY PD
Officer: [Redacted] Inspection: 15.14
Date of Inspection: 06/29/2015
Serial Number: FC-000795
Software: 2100.27

Calibration Test	YES	NO
Temperature/Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol %	0.05g/210L Test (g/210L) Lot#: 2014004 Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 2014004 Exp: 03/25/2016	0.20g/210L Test (g/210L) Lot#: 2014005 Exp: 03/25/2016	0.08 g/210L Dry Gas Bid Test (g/210L) Lot#: 201300025 Exp: 04/01/2015
0.000	0.050	0.061 / 0.080	0.200	0.077
0.000	0.053	0.062 / 0.080	0.200	0.077
0.000	0.051	0.063 / 0.080	0.200	0.077

Inspector: [Redacted]
Agency: Hill County PD, 15.14

The above instrument complies () does not comply () with the Florida Department of Law Enforcement Agency's policy regarding the use and maintenance of this equipment in accordance with the provisions of Chapter 318.02, F.S.

James A. Newcome
Signature and Printed Name

06/29/2015
Date

ORIGINAL

80-000795 (row 4)

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 07 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 11:01

Date of Inspection: 05/03/2015

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 21913080A6 Exp: 09/01/2015
0.000	0.051	0.080	0.200	0.078
0.000	0.051	0.080	0.201	0.078
0.000	0.051	0.081	0.200	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-2, FAC

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Exemption and that I performed this inspection in accordance with the provisions of Chapter 11D-2, FAC.

Russell L. Borch

Signature and Printed Name

RUSSELL L BORCH

05/03/2015
Date

80-000795 (PWS)

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 22 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT -- INOXILYZER 8000

Agency: **FLOR COUNTY SO** State ID Number: **20-000795**
Time of Inspection: **15:38** Date of Inspection: **04/17/2015** Software: **8000 1.1**

Check or Test	YES	NO
Date and/or Time Adjusted		NO
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Alcohol Sus ⁴ Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201405C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201407D Exp: 03/17/2016	0.00 Lot#: 201408E Exp: 03/04/2015
0.000	0.050	0.080	0.199	0.078
0.000	0.051	0.080	0.200	0.077
0.000	0.051	0.080	0.200	0.078

Number of Simulators Used: 0

Remarks:

The above mentioned complies () does not comply () with Chapter 11D-2, F.A.C.

I certify that I am a valid Florida Department of Law Enforcement Agency Inspector and that I performed this inspection in accordance with the provisions of Chapter 11D-2, F.A.C.

Russell T. Burch

Signature and Printed Name

04/17/2015
Date

8U-000795 (S)

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 10:41

Date of Inspection: 03/21/2015

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#: 201305C Exp: 06/21/2015	0.05g/210L Dry Gas Std Test (g/210L) Lot#: 201308026 Exp: 09/31/2015
0.000	0.050	0.081	0.197	0.078
0.000	0.051	0.081	0.196	0.078
0.000	0.051	0.081	0.196	0.078

Number of Calibration Tests: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, F.A.C.

I, the undersigned, am a valid Florida Department of Law Enforcement Agency Inspector and I have performed this inspection in accordance with the provisions of Chapter 11D-8, F.A.C.

Russell A. Birch

Signature and Printed Name

03/21/2015
Date

80-000795 (PWS)

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 30 2015

AGENCY INSPECTION REPORT -- INTOXILYZER Alcohol Testing Program

Agency: GULF COUNTY SO
Time of Inspection: 11:29

Date of Inspection: 02/23/2015

Serial Number: FD-000795
Software: SICG.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201402F Exp: 03/25/2015	0.08g/210L Test (g/210L) Lot#: 201308F Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#: 201402F Exp: 06/25/2015	0.05 g/210L Lot#: 201402F Exp: 03/25/2015
/ 0.000	0.050	0.079	0.195	0.079
/ 0.000	0.050	0.079	0.196	0.078
/ 0.000	0.050	0.080	0.197	0.079

Number of Simulators Used: 5

Remarks

One Ambion K11-BERUN AIRPLANK

The above instrument complies () does not comply () with Chapter 21P-6, F.S.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11P-3, F.A.C.

Russell J. Burke

Signature and Printed Name

RUSSELL J. BURKE

02/23/2015
Date

80-000795 (P)

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 02 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTELLECTUAL PROPERTY

Agency: GUL COUNTY SO
Time of Inspection: 10:33

Date of Inspection: 01/29/2015

Serial Number: 80-000795
Software: 0105.07

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Stop. Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201404F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201404E Exp: 06/22/2016	0.6g/210L Test (g/210L) Lot#: 219130E0AF Exp: 09/01/2016
0.000	0.050	0.078	0.186	0.078
0.000	0.050	0.079	0.197	0.078
0.000	0.051	0.080	0.198	0.078

Inspector: [Signature]

Remarks:

None. ADVANCED TO NCI.

The above instrument complies () does not comply () with Chapter 11F-8, F.S.C.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11F-8, F.S.C.

James A. Johnson

Signature and Printed Name

01/29/2015
Date