88 80-000827

RECEIVED
JAN 04 2016
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Marion County SO
Number: 80-000827
Time of Inspection: 11:00 Date of Inspection: 12/29/2015
Software: 8100.27

Serial

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: ___5_____

Remarks: instrument taken out of service unable to clear chamber gave AMB results.

The above instrument complies (_____) does not comply (__X__) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jeffery Raker
Signature and Printed Name

12/29/2015
Date

FDLE/ATP Form 40 – March 2004

RECEIVED
OCT 29 2015
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 12:23 Date of Inspection: 08/29/2015
Serial Number: 80-000827
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/06/2016	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1669440 Exp: 04/19/2017
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.201	0.080

Number of Simulators Used: 5

Remarks:
WITHIN COMPLIANCE

"AMENDED" .08 LOT # EXPIRATION DATE SHOULD BE 04/08/2016

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JEFFERY A RAKER
Signature and Printed Name

08/29/2015
Date