

Florida Department of Law Enforcement Alcohol Testing Program

DEC 23 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:33

Date of Inspection: 12/19/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081

Number of Simulators Used: 5

Remarks:

A TRUE COPY
 SUSAN BENTON, SHERIFF
 HIGHLANDS COUNTY, FLORIDA
 BY *Shon W. Abeln* 2184 D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

 SHON W ABELN
 Signature and Printed Name

12/19/2015
 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 10:14

Date of Inspection: 11/21/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.047	0.078	0.197	0.079
0.000	0.047	0.077	0.197	0.079
0.000	0.048	0.077	0.196	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY *Shon W. Abeln* 3184 D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

Signature and Printed Name SHON W ABELN

11/21/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 19:18

Date of Inspection: 10/29/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080AZ Exp: 07/05/2016
0.000	0.049	0.079	0.196	0.080
0.000	0.049	0.079	0.195	0.080
0.000	0.049	0.078	0.195	0.080

Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184 SHON W ABELN
Signature and Printed Name

10/29/2015
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

AUG 19 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 14:59

Date of Inspection: 08/16/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.049	0.076	0.194	0.079
0.000	0.049	0.077	0.195	0.079
0.000	0.049	0.077	0.195	0.078

Number of Simulators Used: 5

Remarks:
A F / M A: . I STILL HAD Residual MOUTH Alcohol in my MOUTH From doing It in wrong ORDER. SW

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln D.S.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184

Signature and Printed Name

SHON W ABELN

08/16/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 22 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 12:09

Date of Inspection: 07/18/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2014	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014080A2 Exp: 07/05/2016
0.000	0.047	0.076	0.193	0.081
0.000	0.047	0.076	0.194	0.080
0.000	0.048	0.076	0.195	0.080

Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY *Shon W Abeln* D.S.
2184

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W Abeln

Signature and Printed Name

SHON W ABELN

07/18/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 24 2015

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 21:36

Date of Inspection: 06/20/2015


Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014080A2 Exp: 07/05/2016
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.078	0.198	0.079


Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY  D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

GREGORY S GLASCO

06/20/2015
Date

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

MAY 28 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 10:37

Date of Inspection: 05/23/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.046	0.075	0.191	0.081
0.000	0.046	0.075	0.194	0.079
0.000	0.046	0.076	0.194	0.080

Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY _____ D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gregory S. Glasco
Signature and Printed Name

05/23/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 29 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 01:57

Date of Inspection: 04/25/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014080A2 Exp: 07/05/2016
0.000	0.048	0.076	0.194	0.078
0.000	0.048	0.076	0.194	0.079
0.000	0.047	0.077	0.195	0.078

Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY [Signature] D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

04/25/2015
Date

RECEIVED

Florida Department of Law Enforcement APR 01 2015

Alcohol Testing Program

FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 09:41

Date of Inspection: 03/28/2015

Serial Number: 80-001054
Software: 8100.27

Table with 3 columns: Check or Test, YES, NO. Rows include Date and/or Time Adjusted, Diagnostic Check (Pre-Inspection): OK, Alcohol Free Subject Test: 0.000, Mouth Alcohol Test: Slope Not Met, Interferent Detect Test: Interferent Detect, Diagnostic Check (Post-Inspection): OK.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for three different samples.

Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY [Signature] D.S.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

SHON W ABELN

Signature and Printed Name

03/28/2015
Date

SDN

80-001054

MS

RECEIVED

MAR 02 2015

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 20:05

Date of Inspection: 02/25/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201306G Exp: 06/25/2015	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.047	0.076	0.189 / 0.190	0.079
0.000	0.047	0.076	0.191 / 0.191	0.079
0.000	0.047	0.077	0.192 / 0.191	0.079

Number of Simulators Used: 5

Remarks:
20: Control Outside Tolerance.

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY *Shon W. Abeln* D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln

SHON W ABELN

Signature and Printed Name

02/25/2015
Date

80-001054 (P12) RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JAN 30 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 09:54

Date of Inspection: 01/24/2015

Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201306G Exp: 06/25/2015	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.046	0.076	0.191	0.079
0.000	0.046	0.077	0.192	0.079
0.000	0.046	0.076	0.193	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
SHON W ABELN

Signature and Printed Name

01/24/2015
Date