

80-001119

RECEIVED

OCT 09 2015

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 16:11

Date of Inspection: 10/05/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 301411K Exp: 11/26/2016 | 0.08g/210L Test (g/210L) Lot#: 301404C Exp: 04/08/2016 | 0.20g/210L Test (g/210L) Lot#: 301405B Exp: 08/12/2016 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 3014080AD Exp: 11/05/2016 |
|----------------------------|--|--|--|---|
| 0.000 | 0.051 | 0.080 | 0.196 | 0.075 |
| 0.000 | 0.051 | 0.081 | 0.197 | 0.075 |
| 0.000 | 0.051 | 0.080 | 0.196 | 0.075 |

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 2669/1112
Signature and Printed Name WILLIAM S. MCENNIS

10/05/2015
Date

88 80-00119

RECEIVED

SEP 27 2015

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 12:52

Date of Inspection: 09/24/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201401F Exp: 11/05/2016 | 0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016 | 0.20g/210L Test (g/210L) Lot#: 201408B Exp: 03/10/2016 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 28814030A3 Exp: 11/05/2016 |
|----------------------------|--|--|--|--|
| 0.000 | 0.051 | 0.080 | 0.190 | 0.079 |
| 0.000 | 0.051 | 0.081 | 0.201 | 0.078 |
| 0.000 | 0.051 | 0.081 | 0.201 | 0.079 |

Number of Simulators Used: 5 *(WSA)*

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WES
Signature and Printed Name WILLIAM E. MCKENZIE

09/24/2015
Date

27 So 001119

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

AUG 16 2015

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE
Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 22:04

Date of Inspection: 08/16/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016 | 0.08g/210L Test (g/210L) Lot#: 201404J Exp: 04/08/2016 | 0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 21113080A2 Exp: 12/01/2015 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.078 | 0.200 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.199 | 0.078 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.078 |

Number of Simulators Used: 5

Remarks:

A.P. M. 5:100 NICH MA *JMB*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Susan M. Barce

SUSAN M BARCE

Signature and Printed Name

08/16/2015
Date

80-001119

RECEIVED

JUN 28 2015

FDLE

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:10

Date of Inspection: 06/28/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201401K Exp: 11/25/2015 | 0.08g/210L Test (g/210L) Lot#: 201401C Exp: 04/08/2015 | 0.20g/210L Test (g/210L) Lot#: 201402B Exp: 06/12/2015 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 11110800AL Exp: 12/31/2015 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.079 | 0.201 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.200 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.200 | 0.078 |

Number of Simulators Used: 5

Remarks:

A F / N A: TOO MUCH IN

DMB To much mouth Alcohol.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cpt Susan M Barge

SUSAN M BARGE

Signature and Printed Name

06-28-2015
Date

88 80-001119

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

MAY 14 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:00

Date of Inspection: 05/13/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015 | 0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015 | 0.20g/210L Test (g/210L) Lot#:201306G Exp: 06/25/2015 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31113080A2 Exp: 12/01/2015 |
|----------------------------|---|---|---|---|
| 0.000 | 0.048 | 0.079 | 0.194 | 0.079 |
| 0.000 | 0.048 | 0.079 | 0.196 | 0.079 |
| 0.000 | 0.048 | 0.079 | 0.197 | 0.079 |

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Susan M Barge

SUSAN M BARGE

Signature and Printed Name

05/13/2015
Date

80-001119

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

APR 04 2015

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 19:05

Date of Inspection: 04/04/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201308G Exp: 08/27/2015 | 0.08g/210L Test (g/210L) Lot#: 201308H Exp: 08/27/2015 | 0.20g/210L Test (g/210L) Lot#: 201306G Exp: 08/25/2015 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 3113080AL Exp: 12/01/2015 |
|----------------------------|--|--|--|---|
| 0.000 | 0.048 | 0.079 | 0.195 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.198 | 0.078 |
| 0.000 | 0.048 | 0.079 | 0.198 | 0.078 |

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M. BARDE

Signature and Printed Name

04/04/2015
Date

80-001119

RECEIVED

MAR 06 2015

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 21:37

Date of Inspection: 03/04/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 301306G Exp: 08/27/2015 | 0.08g/210L Test (g/210L) Lot#: 201308H Exp: 08/27/2015 | 0.20g/210L Test (g/210L) Lot#: 701306G Exp: 06/25/2015 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31113080A3 Exp: 12/01/2015 |
|----------------------------|--|--|--|--|
| 0.000 | 0.047 | 0.079 | 0.197 | 0.079 |
| 0.000 | 0.047 | 0.079 | 0.198 | 0.077 |
| 0.000 | 0.048 | 0.079 | 0.197 | 0.078 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cpl. Susan M. Barje

SUSAN M BARJE

Signature and Printed Name

03/04/2015
Date

RECEIVED

FEB 11 2015

Florida Department of Law Enforcement
Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:08

Date of Inspection: 02/09/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015 | 0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015 | 0.20g/210L Test (g/210L) Lot#:201308G Exp: 06/25/2015 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31113080A2 Exp: 12/01/2015 |
|----------------------------|---|---|---|---|
| 0.000 | 0.044 / 0.046 | 0.077 | 0.192 | 0.078 |
| 0.000 | 0.045 / 0.046 | 0.078 | 0.196 | 0.078 |
| 0.000 | 0.046 / 0.047 | 0.078 | 0.198 | 0.078 |

Number of Simulators Used: 5

Remarks:

01: Control Outside Tolerance TIGHTEN LID.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

SUSAN M FARGE

02/09/2015
Date

89 80-001119

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JAN 02 2015

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 18:41

Date of Inspection: 01/01/2015

Serial Number: 80-001119
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201302B Exp: 02/05/2015 | 0.08g/210L Test (g/210L) Lot#: 201302C Exp: 02/05/2015 | 0.20g/210L Test (g/210L) Lot#: 201306G Exp: 06/25/2015 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31113080A2 Exp: 12/01/2015 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.079 | 0.197 | 0.078 |
| 0.000 | 0.049 | 0.080 | 0.198 | 0.078 |
| 0.000 | 0.049 | 0.080 | 0.198 | 0.078 |

Number of Simulators Used: 5

Remarks:
N/A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M. BARGE

Signature and Printed Name

01/01/2015
Date