

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

DEC 14 2015
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 21:21

Date of Inspection: 12/11/2015

Serial Number: 80-001249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG402702 Exp: 01/27/2016
0.000	0.047	0.076	0.200	0.083
0.000	0.048	0.077	0.200	0.083
0.000	0.048	0.078	0.200	0.082

Number of Simulators Used: 5

Remarks:

T. SUN WAS PRESENT & OBSERVED ENTIRE INSPECTION

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo

FRANCA R DICIENZO

Signature and Printed Name

12/11/2015
Date

80-001249

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

NOV 20 2015

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE

Alcohol Testing Program

Agency: BROWARD COUNTY SO
Time of Inspection: 20:31

Date of Inspection: 11/16/2015

Serial Number: 80-001249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG402702 Exp: 01/27/2016
0.000	0.047	0.077	0.200	0.082
0.000	0.048	0.078	0.200	0.082
0.000	0.048	0.077	0.200	0.082


Number of Simulators Used: 5

Remarks:

F. DICIZENZO WAS PRESENT & OBSERVED ENTIRE INSPECTION

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TONY SUN

Signature and Printed Name

11/16/2015
Date

80-001249

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

APR 29 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 20:08

Date of Inspection: 04/27/2015

Serial Number: 80-001249
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG402702 Exp: 01/27/2016
0.000	0.046	0.073 / 0.075	0.194	0.080
0.000	0.047	0.074 / 0.076	0.195	0.079
0.000	0.046	0.075 / 0.076	0.196	0.079

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. ROOM TEMP COOLED SIMULATORS. ANAYA FARZIEE WAS PRESENT.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo

FRANCA R DICIZENZO

Signature and Printed Name

04/27/2015

Date