

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 20 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 19:40

Date of Inspection: 10/19/2015

Serial Number: 80-001250
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016 | 0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016 | 0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1789170 Exp: 01/14/2018 |
|----------------------------|---|---|---|--|
| 0.000 | 0.048 | 0.076 | 0.190 | 0.081 |
| 0.000 | 0.048 | 0.077 | 0.194 | 0.082 |
| 0.000 | 0.048 | 0.078 | 0.195 | 0.082 |

Number of Simulators Used: 5

Remarks:

TONY SUN WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

FRANCA R. DICIZENZO
Signature and Printed Name

10/19/2015
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 20 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 19:00

Date of Inspection: 10/19/2015

Serial Number: 80-001250
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016 | 0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016 | 0.20g/210L Test (g/210L) Lot#: 201402A Exp: 02/11/2016 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp: |
|----------------------------|--|--|--|--|
| 0.000 | 0.048 | 0.078 | 0.199 | |
| 0.000 | 0.048 | 0.079 | 0.198 | |
| 0.000 | 0.049 | 0.078 | 0.199 | |

Number of Simulators Used: 5

Remarks:

TONY SUN WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

Push Enter Button Twice Skipped .08 Dry Gas.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo

FRANCA R DICIENZO

Signature and Printed Name

10/19/2015

Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

AUG 03 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BROWARD COUNTY SO
Time of Inspection: 13:39

Date of Inspection: 07/23/2015

Serial Number: 80-001250
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | | No |
| Alcohol Free Subject Test: 0.000 | | No |
| Mouth Alcohol Test: Slope Not Met | | No |
| Interferent Detect Test: Interferent Detect | | No |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.20g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____ |
|----------------------------|---|---|---|--|
| | | | | |
| | | | | |
| | | | | |

Number of Simulators Used: _____

Remarks:
COMPLIANCE NOT DETERMINED. AI NOT CONDUCTED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

07/23/2015
Date

80-001250

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

APR 29 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BROWARD COUNTY SO
Time of Inspection: 17:07

Date of Inspection: 04/27/2015

Serial Number: 80-001250
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016 | 0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016 | 0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG402702 Exp: 01/27/2016 |
|----------------------------|---|---|---|---|
| 0.000 | 0.047 | 0.077 | 0.198 | 0.082 |
| 0.000 | 0.048 | 0.077 | 0.199 | 0.082 |
| 0.000 | 0.048 | 0.077 | 0.199 | 0.082 |

Number of Simulators Used: 5

Remarks:
TONY SUN WAS PRESENT AND OBSERVED ENTIRE INSPECTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

FRANCA R. DICIZENZO
Signature and Printed Name

04/27/2015
Date