

RECEIVED

NOV 06 2015

Florida Department of Law Enforcement

Alcohol Testing Program *JGM*

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MELBOURNE P.D.

Time of Inspection: 12:32

Date of Inspection: 11/03/2015

Serial Number: 80-001262

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1841081 Exp: 04/08/2018
0.000	0.050	0.079	0.202	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.079	0.201	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JGM
516
404

MARK D WHITRIGHT

Signature and Printed Name

Chad S. Lewis

11/03/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

NOV 06 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: MELBOURNE P.D.

Time of Inspection: 11:51

Date of Inspection: 11/03/2015

Serial Number: 80-001262

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: 4

Remarks:

A F / M A: Interferent Detect. Non-compliance: *FIRST ATTEMPT: INTERFERENT DETECT. SECOND ATTEMPT: REGISTER BREATH RESULT OF 0.113G/210L. DISCONTINUED AGENCY INSPECTION AND INITIATED A NEW INSPECTION.*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


MARK D WHITRIGHT

 Signature and Printed Name

11/03/2015
 Date

Chris S Livingston