

80-001265
②

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Serial Number: 80-001265

Time of Inspection: 09:44

Date of Inspection: 03/28/2015

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016 | 0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015 | 0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32114080A2 Exp: 12/05/2016 |
|----------------------------|---|---|---|---|
| 0.000 | 0.048 | 0.078 | 0.196 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.081 |
| 0.000 | 0.049 | 0.080 | 0.199 | 0.080 |


Number of Simulators Used: 4

Remarks:

A F / M A: Range Exceeded OPER ERROR. Int Det:FOGOT TO CONNECT.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



D. MORGAN

Signature and Printed Name

03/28/2015
Date

9

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 10:14

Date of Inspection: 02/26/2015

Serial Number: 80-001265
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016 | 0.08g/210L Test (g/210L) Lot#: 201308H Exp: 08/27/2015 | 0.20g/210L Test (g/210L) Lot#: 201402A Exp: 02/11/2016 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 25413080A2 Exp: 10/01/2015 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.078 | 0.197 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.080 |


Number of Simulators Used: 4

Remarks:

Int Det: AIR LEAK.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DARIN MORGAN

Signature and Printed Name

02/26/2015
Date

10.22 01/25/2015

PAUL BRY P.O.
Intoxilyzer - Alcohol Analyzer
Model: 8000
01/25/2015
Software: 8100.27

| Test | 9/21/0L | Time |
|----------------|---------|-------|
| Air Blank | 0.000 | 12:19 |
| Control Test | 0.049 | 12:19 |
| Air Blank | 0.000 | 12:20 |
| Control Test | 0.049 | 12:20 |
| Air Blank | 0.000 | 12:21 |
| Control Test | 0.049 | 12:21 |
| Air Blank | 0.000 | 12:22 |
| Control Test | 0.049 | 12:22 |
| Average | 0.0499 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

Operator's Signature

PAUL BRY P.O.
Intoxilyzer - Alcohol Analyzer
Model: 8000
01/25/2015
Software: 8100.27

| Test | 9/21/0L | Time |
|----------------|---------|-------|
| Air Blank | 0.000 | 12:23 |
| Control Test | 0.077 | 12:24 |
| Air Blank | 0.000 | 12:25 |
| Control Test | 0.075 | 12:25 |
| Air Blank | 0.000 | 12:26 |
| Control Test | 0.078 | 12:27 |
| Air Blank | 0.000 | 12:27 |
| Control Test | 0.078 | 12:27 |
| Average | 0.0780 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.2821 | |

Operator's Signature

PAUL BRY P.O.
Intoxilyzer - Alcohol Analyzer
Model: 8000
01/25/2015
Software: 8100.27

| Test | 9/21/0L | Time |
|----------------|---------|-------|
| Air Blank | 0.000 | 12:08 |
| Control Test | 0.194 | 12:08 |
| Air Blank | 0.000 | 12:09 |
| Control Test | 0.196 | 12:10 |
| Air Blank | 0.000 | 12:10 |
| Control Test | 0.197 | 12:11 |
| Air Blank | 0.000 | 12:11 |
| Control Test | 0.1957 | 12:11 |
| Average | 0.1957 | |
| Std Dev | 0.0015 | |
| Rel Std Dev(%) | 0.7697 | |

Operator's Signature

D. Morgan 2-26-2015