

RECEIVED
80-001286
DEC 08 2015

Florida Department of Law Enforcement

Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 20:39

Date of Inspection: 11/29/2015

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02814080A3 Exp: 03/01/2016
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.078	0.201	0.078
0.000	0.049	0.079	0.201	0.078

Number of Simulators Used: 5

Remarks:


BAKER COUNTY
SHERIFF'S OFFICE

DEC 08 2015

RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



COREY S WILSON

Signature and Printed Name

11/29/2015
Date

RECEIVED

89 80-001286
DEC 08 2015

Florida Department of Law Enforcement

Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 11:34

Date of Inspection: 10/31/2015

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02814080A3 Exp: 03/01/2016
0.000	0.048	0.078	0.201	0.079
0.000	0.047	0.079	0.201	0.079
0.000	0.048	0.079	0.201	0.078

Number of Simulators Used: 5


Remarks:
Time-Date changed.

BAKER COUNTY
SHERIFF'S OFFICE
DEC 03 2015

RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 _____ COREY S WILSON
 Signature and Printed Name

10/31/2015
 Date

80-001286 BK

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 19 2015

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 03:10

Date of Inspection: 09/30/2015

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02614080A3 Exp: 03/01/2016
0.000	0.048	0.076	0.198	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.200	0.078

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE

OCT 12 2015

RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

09/30/2015
Date

Florida Department of Law Enforcement
 Alcohol Testing Program

AUG 04 2015

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
 Time of Inspection: 16:06

Date of Inspection: 07/29/2015

Serial Number: 80-001286
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02E14080A3 Exp: 03/01/2016
0.000	0.048	0.077	0.199	0.078
0.000	0.047	0.078	0.199	0.078
0.000	0.048	0.078	0.200	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson 263 _____ COREY S WILSON
 Signature and Printed Name

07/29/2015
 Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

AUG 04 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 02:14

Date of Inspection: 06/30/2015

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02814080A3 Exp: 03/01/2016
0.000	0.046	0.075	0.197	0.078
0.000	0.047	0.076	0.197	0.078
0.000	0.047	0.076	0.197	0.078

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE

JUL 30 2015

RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

06/30/2015
Date

88-001286
RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 03 2015

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 07:38

Date of Inspection: 05/28/2015

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201306G Exp: 06/25/2015	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02814080A3 Exp: 03/01/2016
0.000	0.047	0.077	0.197	0.079
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.078	0.198	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

COREY S WILSON

05/28/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001286
RECEIVED
MAY 08 2015
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 13:14

Date of Inspection: 04/26/2015

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201306G Exp: 06/25/2015	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02B14080A3 Exp: 03/01/2016
0.000	0.047	0.078	0.198	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.078	0.199	0.078

Number of Simulators Used: 5


Remarks:

BAKER COUNTY
SHERIFF'S OFFICE
MAY 04 2015

RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

COREY S WILSON

04/26/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001286
RECEIVED
MAY 08 2015
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 22:17

Date of Inspection: 03/28/2015

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201308G Exp: 08/27/2015	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201306G Exp: 06/25/2015	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02814080A3 Exp: 03/01/2016
0.000	0.048	0.079	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.048	0.080	0.201	0.079

Number of Simulators Used: 5

Remarks:

**BAKER COUNTY
SHERIFF'S OFFICE**
MAY 04 2015
**RECORDS SECTION
RECEIVED**

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson 263 _____
Signature and Printed Name

03/28/2015
Date