

# Florida Department of Law Enforcement Alcohol Testing Program *PJM*

**RECEIVED**

AUG 04 2015

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SANTA ROSA COUNTY SO  
Time of Inspection: 08:21

Date of Inspection: 08/04/2015

Serial Number: 80-001302  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
SKIPPED AI TO OPERATE INSTRUMENT; NOT A COMPLIANCE CHECK

The above instrument complies ( *N/A* ) does not comply ( *PHS* ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHILLIP W SUBER

Signature and Printed Name

08/04/2015  
Date

# Florida Department of Law Enforcement

## Alcohol Testing Program

**RECEIVED**

AUG 04 2015

#13LE

### AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: SANTA ROSA COUNTY SO  
 Time of Inspection: 11:00

Date of Inspection: 05/28/2015

Serial Number: 80-001302  
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.050	0.078	0.199	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079


Number of Simulators Used: 5

Remarks:

COPY

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

DURON A NELSON

05/28/2015  
 Date