

80-001321
①

RECEIVED

Florida Department of Law Enforcement SEP 15 2015 Alcohol Testing Program FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO
Time of Inspection: 13:49

Date of Inspection: 09/15/2015

Serial Number: 80-001321
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1686121 Exp: 05/21/2017
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.079	0.200	0.081

Number of Simulators Used: 4

Remarks:

AF / MA: .

On the initial mouth alcohol test, I did not provide a sufficient sample. I repeated the AF/MA test and it was successful.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MARK W GARLAND

Signature and Printed Name

09/15/2015
Date