



1416

Sheriff Jerry L. Demings

# ORANGE COUNTY SHERIFF'S OFFICE

TELEPHONE: 407.254.7000 ❖ P.O. BOX 1440, ORLANDO, FLORIDA 32802-1440 ❖ WWW.OCSO.COM

RECEIVED

DEC 14 2015

FDLE  
Alcohol Testing Program

December 8, 2015

Department Inspector Jake Shanahan  
Florida Department of Law Enforcement  
Alcohol Testing Program  
P.O. Box 1489  
Tallahassee, FL 32302-1489

Dear Inspector Shanahan:

This letter is to recap our telephone conversation on December 8, 2015 regarding the incident which occurred during the Agency Inspection conducted on December 8, 2015.

The Agency Inspection of Intoxilyzer 8000, serial number 80-001416, was interrupted due to an unknown power issue. The instrument was in the Agency Inspection mode, and was in the process of conducting the air blank after the Alcohol Free/Mouth Alcohol test. The instrument had complied with all of the Diagnostic Checks, and had successfully completed the Alcohol Free/Mouth Alcohol test. During the air blank, the instrument's front display went blank, and the green power indicator light on the front of the instrument switched color to a red indicator light. I had not touched anything on the instrument, moved the instrument, or in any way caused this to occur.

I pushed the green "Start Test" button to power the instrument back on. I verified that the power cord was properly plugged into the instrument, which I found to be the case. I went to the "Recall" function in the Agency Inspector menu, and found that there was no record of the partial Agency Inspection which had been interrupted by the unknown power issue.

I began another Agency Inspection later in the morning on instrument 80-001416. The instrument successfully completed the Agency Inspection, and complied with Rule 11D-8. I notified you of this event via telephone.



*The First Agency in Orange County to Receive Both  
International and State Accreditation.*



1416  
②

Department Inspector Jake Shanahan  
December 8, 2015  
Page 2

**RECEIVED**


DEC 14 2015

**FDLE  
Alcohol Testing Program**

If you have any questions, please contact me at (407) 836-0800.

Sincerely,

Jerry L. Demings  
Sheriff of Orange County



Agency Inspector Kelly Melville  
Special Operations Division  
Traffic/DUI Testing Center

KM/kmm