

# Florida Department of Law Enforcement

## Alcohol Testing Program *ggm*

*AK*

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO  
 Time of Inspection: 17:32

Date of Inspection: 09/28/2015

Serial Number: 80-001439  
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34513080A7 Exp: 01/01/2016
0.000	0.050	0.077	0.202	0.077
0.000	0.050	0.077	0.202	0.077
0.000	0.050	0.078	0.202	0.077

Number of Simulators Used: 5

Remarks:

ANNUAL DEPT. INSPECTION

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Bradford B Baker*

BRADFORD B BAKER

---

Signature and Printed Name

09/28/2015  
Date