

80-001721
②

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Serial Number: 80-001721

Time of Inspection: 10:00

Date of Inspection: 03/28/2015

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#: 201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32114080A2 Exp: 12/05/2016
0.000	0.050	0.079	0.197	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.050	0.079	0.197	0.079


Number of Simulators Used: 4

Remarks:

Time-Date changed.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



D. MORGAN

Signature and Printed Name

03/28/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000



Agency: PALM BAY P.D.
Time of Inspection: 09:48

Date of Inspection: 02/26/2015

Serial Number: 80-001721
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#: 201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 25413050A2 Exp: 10/01/2015
0.000	0.051	0.079	0.195	0.079
0.000	0.050	0.079	0.196	0.079
0.000	0.050	0.079	0.196	0.079

Number of Simulators Used: 4

Remarks:
Time-Date changed.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



D. MORGAN

Signature and Printed Name

02/26/2015
Date