

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

NOV 20 2015

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Serial Number: 80-001737

Software: 8100.27

Agency: SOUTH MIAMI PD

Time of Inspection: 07:39

Date of Inspection: 10/29/2015

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | | No |
| Alcohol Free Subject Test: 0.000 | | No |
| Mouth Alcohol Test: Slope Not Met | | No |
| Interferent Detect Test: Interferent Detect | | No |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.20g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____ |
|----------------------------|---|---|---|--|
| | | | | |
| | | | | |
| | | | | |


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED. AI NOT CONDUCTED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

10/29/2015
Date