80-002988

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI GARDENS PD
Time of Inspection: 14:53 Date of Inspection: 07/28/2015
Serial Number: 80-002988
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24813080A3 Exp: 10/01/2015
0.000	0.050	0.080	0.206	0.079
0.000	0.050	0.080	0.206	0.079
0.000	0.050	0.080	0.206	0.080

Number of Simulators Used: 5

Remarks:
IN COMPLIANCE

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BRIAN P OMALLEY
Signature and Printed Name

07/28/2015
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI GARDENS PD
Serial Number: 80-002988
Time of Inspection: 08:58
Date of Inspection: 02/25/2015
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24813080A3 Exp: 10/01/2015
0.000	0.047	0.081	0.208	0.080
0.000	0.048	0.081	0.208	0.079
0.000	0.048	0.081	0.208	0.080

Number of Simulators Used: 5

Remarks:
IN COMPLIANCE

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

#217

BRIAN P OMALLEY
Signature and Printed Name

02/25/2015
Date

FDLE/ATP Form 40 – March 2004