

80-003407

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

DEC 17 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 22:25

Date of Inspection: 12/17/2015

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201509A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 2015000 Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201509A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 28814080AS Exp: 11/05/2016
0.000	0.050	0.080	0.195	0.079
0.000	0.050	0.080	0.196	0.078
0.000	0.050	0.080	0.197	0.078

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

12/17/2015
Date

89 80-003407

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

NOV 11 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 18:26

Date of Inspection: 11/11/2015

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 2014041Y Exp: 11/25/2015	0.08g/210L Test (g/210L) Lot#: 2014040C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201402A Exp: 02/01/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31113080A2 Exp: 10/01/2015
0.000	0.050	0.079	0.195	0.078
0.000	0.050	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.078

Number of Simulators Used: 0

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

11/11/2015
Date

89 80-003407

RECEIVED

Florida Department of Law Enforcement OCT 09 2015
Alcohol Testing Program

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:48

Date of Inspection: 10/05/2015

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201406F Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201406G Exp: 04/09/2016	0.20g/210L Test (g/210L) Lot#: 201408E Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31113080A2 Exp: 12/01/2015
0.000	0.049	0.078	0.199	0.077
0.000	0.049	0.078	0.199	0.077
0.000	0.049	0.078	0.199	0.077

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 2669/1112

Signature and Printed Name WILLIAM S MCNEELIE

10/05/2015
Date

80-003407

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

OCT 09 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:48

Date of Inspection: 10/05/2015

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411F Exp: 11/29/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201408E Exp: 05/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 311-3000A2 Exp: 12/01/2015
0.000	0.049	0.078	0.199	0.077
0.000	0.049	0.078	0.199	0.077
0.000	0.049	0.078	0.199	0.077

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 2669/1112

WILLIAM G. MCENENIE

Signature and Printed Name

10/05/2015
Date

8980-003407

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

SEP 27 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 12:31

Date of Inspection: 09/24/2015

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:301411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:301404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:301405B Exp: 02/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31113080A2 Exp: 12/01/2015
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.048	0.080	0.197	0.078

Number of Simulators Used: 2

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SUSAN M. BARGE

Signature and Printed Name

09/24/2015
Date

89 80-003407

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

AUG 16 2015

AGENCY INSPECTION REPORT - INTOXILYZER 8000 FDLE Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 21:51

Date of Inspection: 08/16/2015

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 0011040 Exp: 01/25/2016	0.08g/210L Test (g/210L) Lot#: 0011040 Exp: 01/25/2016	0.20g/210L Test (g/210L) Lot#: 0011040 Exp: 01/25/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1113080A3 Exp: 12/01/2015
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.078	0.196	0.078

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M. BARGE

Signature and Printed Name

08/16/2015
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 06 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 16:25

Date of Inspection: 07/06/2015

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201406E Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 31113089A2 Exp: 12/01/2015
0.000	0.050	0.079	0.199	0.078
0.000	0.051	0.079	0.199	0.076
0.000	0.050	0.078	0.199	0.078

Number of Simulators Used: 5

Remarks:
N/A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name JAYON D. MURPHY

07/06/2015
Date

80-003407

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

JUN 28 2015

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-003407

Time of Inspection: 10:33

Date of Inspection: 06/28/2015

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201404K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1111089A7 Exp: 12/31/2015
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.078

Number of Simulators Used: 0

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAYON D. MURREY

Signature and Printed Name

06/28/2015
Date

MAR 11 2015

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

3/11/15
003407

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 21:22

Date of Inspection: 03/04/2015

Serial Number: -80-000000-
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201308G Exp: 08/27/2015	0.08g/210L Test (g/210L) Lot#: 201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#: 201306G Exp: 05/25/2015	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014080A2 Exp: 07/05/2016
0.000	0.048	0.077	0.196	0.080
0.000	0.047	0.078	0.195	0.079
0.000	0.048	0.078	0.195	0.079

Number of Simulators Used: 5

Remarks:
N/A

The serial number for this instrument was erased and the form has been revised with the serial number 003407. According to State Agency Inspector, Jennifer Jacobs, this will not interfere with the function of the instrument.

[Signature] 3/11/15

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

JAVON D MURPHY
Signature and Printed Name

03/04/2015
Date

88 JB 003407

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JAN 02 2015

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 18:12

Date of Inspection: 01/02/2015

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201307B Exp: 02/05/2015	0.08g/210L Test (g/210L) Lot#: 201302C Exp: 02/05/2015	0.20g/210L Test (g/210L) Lot#: 201306B Exp: 06/25/2015	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1601408052 Exp: 07/05/2016
0.000	0.048	0.078	0.195	0.079
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.197	0.079

Number of Simulators Used: 5

Remarks:
N/A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN K. BARGE

Signature and Printed Name

01/02/2015
Date