

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUN 22 2015
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 09:15

Date of Inspection: 06/19/2015

Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046	0.073 / 0.075	0.187 / 0.188	
0.000	0.046	0.074 / 0.076	0.190 / 0.189	
0.000	0.046	0.074 / 0.075	0.190 / 0.189	

Number of Simulators Used: 5

Remarks:

RETIGHTENED SEALS AND RESTARTED TEST. RETIGHTENED SEAL AND RESTRTED TEST, Control Outside Toleran. F AILED TO REACH LEVELS ON .20 TWICE.

RECEIVED
JUN 22 2015
FDLE
Alcohol Testing Program

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

GREGORY DOUBERLEY

06/19/2015
Date