

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

**RECEIVED**  
SEP 28 2015  
FDLE  
Alcohol Testing Program

Agency: LAKELAND PD  
Time of Inspection: 10:13

Date of Inspection: 09/23/2015

Serial Number: 80-005810  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:27713080A1 Exp: 11/01/2015
0.000	0.048	0.079	0.196	0.078
0.000	0.049	0.080	0.197	0.078
0.000	0.049	0.080	0.198	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz \_\_\_\_\_ GORDON L FRANZ  
Signature and Printed Name

09/23/2015  
Date

80-005810 *DN*

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
AUG 12 2015  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 10:23

Date of Inspection: 08/10/2015

Serial Number: 80-005810  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:27713080A1 Exp: 11/01/2015
0.000	0.049	0.078	0.196	0.078
0.000	0.049	0.078	0.197	0.079
0.000	0.050	0.078	0.197	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ  
Signature and Printed Name

08/10/2015  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 07 2015

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: LAKELAND PD  
Time of Inspection: 12:11

Date of Inspection: 07/07/2015

Serial Number: 80-005810  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:27713080A1 Exp: 11/01/2015
0.000	0.048	0.079	0.198	0.077
0.000	0.048	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ  
Signature and Printed Name

07/07/2015  
Date

RECEIVED  
MAY 19 2015  
FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 11:45

Date of Inspection: 05/18/2015

Serial Number: 80-005810  
Software: 8100.27.00

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#:201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:27713080A1 Exp: 11/01/2015
0.000	0.047 / 0.048	0.077	0.197	0.081
0.000	0.047 / 0.048	0.077	0.199	0.082
0.000	RFI / 0.049	0.078	0.199	0.082

Number of Simulators Used: 5

Remarks:

05: RFI Detect.

.05 TEST REPEATED AND PASSED *G.L. Franz*

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz GORDON L FRANZ  
Signature and Printed Name

05/18/2015  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 09 2015

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 13:27

Date of Inspection: 04/09/2015

Serial Number: 80-005810  
Software: 8100.27.00

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201308H Exp: 08/27/2015	0.20g/210L Test (g/210L) Lot#: 201402A Exp: 02/11/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 27713080A1 Exp: 11/01/2015
0.000	0.049	0.078	0.199	0.083
0.000	0.049	0.079	0.199	0.082
0.000	0.049	0.079	0.200	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GORDON L FRANZ  
Signature and Printed Name

04/09/2015  
Date