

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

SEP 02 2015

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 08:06

Date of Inspection: 08/29/2015

Serial Number: 80-005935
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046	0.074 / 0.074		
0.000	0.046	0.075 / 0.075		
0.000	0.046	0.075 / 0.075		


Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance/CHECKED FITTINGS, Control Out. Non-compliance: .08 READING TOO LOW.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

08/29/2015
 Date