RECEIVED ✓

SEP 17 2015

FDLE
Alcohol Testing Program

# Florida Department of Law Enforcement
# Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 16:49
Date of Inspection: 09/16/2015
Serial Number: 80-000833
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|---|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

✓ BK

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016 ✓ | 0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017 ✓ | 0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016 ✓ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG511701 Exp: 04/27/2017 ✓ |
|---|---|---|---|---|
| 0.000 | 0.050 | 0.080 | 0.202 | 0.077 |
| 0.000 | 0.050 | 0.080 | 0.203 | 0.077 |
| 0.000 | 0.050 | 0.080 | 0.203 | 0.077 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.078 |
| 0.000 | 0.050 | 0.080 | 0.203 | 0.077 |
| 0.000 | 0.050 | 0.081 | 0.203 | 0.078 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.078 |
| 0.000 | 0.050 | 0.081 | 0.204 | 0.077 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.077 |
| 0.000 | 0.050 | 0.080 | 0.204 | 0.077 |

| Standard Deviations | 0.0004 | 0.0005 | 0.0006 | 0.0004 |
|---|---|---|---|---|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIELLE S MILLER
Signature and Printed Name

09/16/2015
Date

*FDLE/ATP Form 41 –Revised August 2005*

RECEIVED
JUN 26 2015
FDLE
Alcohol Testing Program

# Florida Department of Law Enforcement
# Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 13:39
Date of Inspection: 06/25/2015
Serial Number: 80-000833
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|---|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

✓BK

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016 ✓ | 0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017 ✓ | 0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016 ✓ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG502701 Exp: 01/27/2017 ✓ |
|---|---|---|---|---|
| 0.000 | 0.050 | 0.080 | 0.200 | 0.078 |
| 0.000 | 0.050 | 0.080 | 0.201 | 0.077 |
| 0.000 | 0.051 | 0.081 | 0.200 | 0.078 |
| 0.000 | 0.051 | 0.081 | 0.201 | 0.077 |
| 0.000 | 0.051 | 0.081 | 0.201 | 0.077 |
| 0.000 | 0.050 | 0.081 | 0.201 | 0.078 |
| 0.000 | 0.051 | 0.080 | 0.201 | 0.077 |
| 0.000 | 0.051 | 0.081 | 0.201 | 0.077 |
| 0.000 | 0.051 | 0.081 | 0.201 | 0.077 |
| 0.000 | 0.051 | 0.081 | 0.201 | 0.076 |
| Standard Deviations | 0.0004 | 0.0004 | 0.0004 | 0.0006 |

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

99

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle S Miller
DANIELLE S MILLER
Signature and Printed Name

06/25/2015
Date

*FDLE/ATP Form 41 –Revised August 2005*