# Florida Department of Law Enforcement
# Alcohol Testing Program

RECEIVED
DEC 29 2015
FDLE
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Serial Number: 80-000869
Time of Inspection: 10:32
Date of Inspection: 12/17/2015
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|---|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017 | 0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG431502 Exp: 11/11/2016 |
|---|---|---|---|---|
| 0.000 | 0.052 | 0.081 | 0.203 | 0.081 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.079 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.079 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.079 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.081 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.080 |
| 0.000 | 0.052 | 0.081 | 0.204 | 0.080 |

| Standard Deviations | 0.0004 | 0.0000 | 0.0005 | 0.0007 |
|---|---|---|---|---|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Eduardo Reyes Rivera
Signature and Printed Name

DAVID E REYES-RIVERA

12/17/2015
Date

*FDLE/ATP Form 41 –Revised August 2005*