Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
SEP 09 2015
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GLADES COUNTY S.O.
Time of Inspection: 14:34
Date of Inspection: 09/08/2015
Serial Number: 80-000947
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

BK

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016 ✓	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017 ✓	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016 ✓	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG511701 Exp: 04/27/2017 ✓
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.080	0.200	0.078

Standard Deviations	0.0003	0.0004	0.0005	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIELLE S MILLER
Signature and Printed Name

09/08/2015
Date