RECEIVED ✓

MAY 02 2015

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 10:14

Date of Inspection: 05/01/2015

Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429701 Exp: 10/24/2016
0.000	0.049	0.080	0.200	0.082
0.000	0.049	0.080	0.200	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.081	0.201	0.081

Standard Deviations	0.0004	0.0003	0.0004	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIELLE S MILLER
Signature and Printed Name

05/01/2015
Date

FDLE/ATP Form 41 –Revised August 2005

QC-SDN