RECEIVED ✓

MAR 26 2015

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 14:51
Date of Inspection: 03/25/2015
Serial Number: 80-001238
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

✓ BK

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016 ✓	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017 ✓	0.20g/210L Test (g/210L) Lot#:201402A Exp: 02/11/2016 ✓	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG421103 Exp: 07/30/2016 ✓
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.079

Standard Deviations	0.0004	0.0004	0.0003	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle S Miller — DANIELLE S MILLER
Signature and Printed Name

03/25/2015
Date

FDLE/ATP Form 41 –Revised August 2005