RECEIVED
OCT 05 2015
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FWC
Time of Inspection: 11:08 Date of Inspection: 09/09/2015
Serial Number: 80-005248
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

BK

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG431502 Exp: 11/11/2016
0.000	0.050	0.000 / 0.080	0.199	0.080
0.000	0.051	0.000 / 0.080	0.199	0.080
0.000	0.050	0.000 / 0.080	0.198	0.080
0.000	0.051	0.000 / 0.081	0.199	0.079
0.000	0.051	0.000 / 0.081	0.199	0.079
0.000	0.051	0.000 / 0.080	0.199	0.079
0.000	0.050	0.000 / 0.081	0.199	0.079
0.000	0.051	0.000 / 0.081	0.199	0.079
0.000	0.051	0.000 / 0.081	0.200	0.079
0.000	0.051	0.000 / 0.080	0.200	0.079

Standard Deviations	0.0004	0.0000 / 0.0005	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:
SIMULATOR NOT CONNECTED.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DAVID E REYES-RIVERA
Signature and Printed Name

09/09/2015
Date

FDLE/ATP Form 41 –Revised August 2005