RECEIVED

FEB 25 2015

FDLE
Alcohol Testing Program

# Florida Department of Law Enforcement
# Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 12:46

Date of Inspection: 02/18/2015

Serial Number: 80-006261
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|---|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

BK

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2016 ✓ | 0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016 ✓ | 0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016 ✓ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG421103 Exp: 07/30/2016 ✓ |
|---|---|---|---|---|
| 0.000 | 0.050 | 0.081 | 0.201 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.205 | 0.081 |
| 0.000 | 0.052 | 0.083 | 0.205 | 0.081 |
| 0.000 | 0.052 | 0.082 | 0.205 | 0.080 |
| 0.000 | 0.051 | 0.083 | 0.205 | 0.081 |
| 0.000 | 0.052 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.052 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.052 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.052 | 0.083 | 0.204 | 0.081 |
| Standard Deviations | 0.0007 | 0.0006 | 0.0011 | 0.0003 |

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PHILLIP W SUBER
Signature and Printed Name

02/18/2015
Date

*FDLE/ATP Form 41 –Revised August 2005*