FDLE
Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Manatee County SO Instrument Serial Number: 80-001742

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
☐ Registration ☐ Request for Registration Attached ☐ Annual ☐ Inspection After Repair ☐ Other: _____ ☐ Barometric Pressure: _____ Gauge _____ Instrument ☐ Instrument Set Up Verified	☐ Agency Inspection Notification ☐ Agency Contact ☐ Other _____ Date of Notification/Contact: 12/7/2015 Agency Inspector: Dep Cline	The following records were audited: ☐ Agency Inspection Reports ☐ Breath Test Instrument Repair/Maintenance Records ☐ Instrument Registration ☐ Dry Gas Standard Certificate(s) of Analysis ☐ Other: _____ ☐ Other: _____

Comments:
Met with Dep Cline and Sgt Spurlock at jail facility. Breath test room was clean and access was limited. Instrument was further secured in a locked cabinet.

Instrument/Area:	Equipment:	Supplies:
☒ Clean/Dry ☒ Secure ☒ Limited Access ☐ Other: _____	☐ Proper Number of Simulators ☐ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☐ Class A Glassware	☐ Distilled/Deionized Water ☐ Mouth Alcohol Solution ☐ Acetone Stock Solution ☐ Alcohol Reference Solution ☐ Dry Gas Standard ☐ Mouth Pieces

Comments:
Suggested that instrument cabinet have better venting if lower control tests persists.

Action:	Comments:
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC	Accompanied by FDLE Dept. Inspector David Reyes.
☒ Remain in Evidentiary Use ☐ Return to/Place into Evidentiary Use ☐ Remove from Evidentiary Use ☐ Remain Out of Evidentiary Use	
☐ Conduct an Agency Inspection Before Evidentiary Use	

[signature] 12/7/15
Signature of Alcohol Testing Program Staff Member Date

Original – FDLE Copy – Agency