FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Osceola County SO S/N 80-000968

Date In 10/27/15 Date Out 11/10/15 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED NOV 10 2015 FDLE Alcohol Testing Program

Intake Performed By TP

- ☐ Registration
- ☑ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:

- ☑ Power cord
- ☐ Printer Cable
- ☐ Other

Notes:

Quality Checks Performed By

- ☑ Breath Tube Screen
- ☑ Replace O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 194
- ☑ Flow Verification (L/s)
  - Flow Column # APIC5
  - 32mm 152 (.139 - .169)
  - 36mm 177 (.156 - .190)
  - 53mm 253 (.228 - .278)
  - 103mm 508 (.447 - .547)
- ☑ Barometric Pressure Check
  - Gauge ID # 26932
- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1024	2014 11K / 11/25/16
0.08	DR1279	2015 02G / 02/24/17
0.20	DR3856	2014 08B / 08/12/16
0.08 DGS	N/A	AGS1174 / 4/27/17

Flow Calibration Performed By

- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete
  - Flow Column #
  - ☐ 5L/min – 17mm
  - ☐ 15L/min – 53mm
  - ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)
  - Flow Column #
  - 32mm (.139 - .169)
  - 36mm (.156 - .190)
  - 53mm (.228 - .278)
  - 103mm (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☑ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By

- ☑ Barometric Pressure ID# 26932
  - 1017 Gauge
  - 1017 Instrument

Mouth Alcohol Solution Lot # 2015-A

Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1016
Interferent	G2840
0.05	SD1024
0.08	DR1279
0.20	DR3856

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: QC - BK

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

Patrick Murphy
Quality Control Review

11/10/15
Date

Revised November 2012

80-000968
Stability checks
11/10/15

BGM

INTOXILYZER 8000
Instrument Initialization
08:16 11/10/2015

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
11/10/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:50
Control Test	0.050	08:51
Air Blank	0.000	08:51
Control Test	0.050	08:52
Air Blank	0.000	08:53
Control Test	0.050	08:53
Air Blank	0.000	08:54
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
11/10/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:57
Control Test	0.080	08:57
Air Blank	0.000	08:57
Control Test	0.079	08:58
Air Blank	0.000	08:58
Control Test	0.080	08:59
Air Blank	0.000	08:59
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
11/10/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:01
Control Test	0.202	09:02
Air Blank	0.000	09:02
Control Test	0.203	09:03
Air Blank	0.000	09:03
Control Test	0.203	09:04
Air Blank	0.000	09:05
Control Test Stats		
Average	0.2027	
Std Dev	0.0006	
Rel Std Dev(%)	0.2849	

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
11/10/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:22
Control Test	0.078	09:22
Air Blank	0.000	09:23
Control Test	0.079	09:23
Air Blank	0.000	09:24
Control Test	0.079	09:25
Air Blank	0.000	09:25
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	BK
Rel Std Dev(%)	0.7339	

Operator's Signature

FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Osceola Co SO S/N 80-000968

Date In 5/19/15 Date Out 5/27/15 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED MAY 27 2015 FDLE Alcohol Testing Program

Intake Performed By

- ☐ Registration
- ☒ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other

Visual Inspection:

- OK Case
- OK Handle
- OK Dry Gas Holder
- OK Feet
- OK Keyboard/Plug
- OK Back/Plugs
- OK Screws tight
- OK Breath Hose

Other Equipment:

- ☐ Power cord
- ☐ Printer Cable
- ☐ Other

Notes:

Quality Checks Performed By

- ☒ Breath Tube Screen
- ☒ Replace O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 197 ✓
- ☒ Flow Verification (L/s)
  - Flow Column # ATP101
  - 32mm 160 (.139 - .169) ✓
  - 36mm 174 (.156 - .190) ✓
  - 53mm 243 (.228 - .278) ✓
  - 103mm 502 (.447 - .547) ✓
- ☒ Barometric Pressure Check
  - Gauge ID # 28427
- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201411K 11/25/16
0.08	SD1011	2015026 2/24/17
0.20	SD1025	2014088 8/12/16
0.08 DGS	N/A	AG502603 01/26/17

Flow Calibration Performed By

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete
  - Flow Column #
  - ☐ 5L/min – 17mm
  - ☐ 15L/min – 53mm
  - ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)
  - Flow Column #
  - 32mm (.139 - .169)
  - 36mm (.156 - .190)
  - 53mm (.228 - .278)
  - 103mm (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By

☒ Barometric Pressure ID# 28427 — 1021 Gauge, 1020 Instrument

Mouth Alcohol Solution Lot # 2015-A

Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

- ☒ Form 41
- ☒ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: OK - BK

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

Patrick Murphy
Quality Control Review

5/27/15
Date

Revised November 2012

80-000968
Stability Checks
5/27/15

INTOXILYZER 8000
Instrument Initialization
07:29 05/27/2015

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
05/27/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:15
Control Test	0.048	08:15
Air Blank	0.000	08:16
Control Test	0.048	08:16
Air Blank	0.000	08:17
Control Test	0.048	08:18
Air Blank	0.000	08:18
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
05/27/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:19
Control Test	0.077	08:20
Air Blank	0.000	08:20
Control Test	0.077	08:21
Air Blank	0.000	08:21
Control Test	0.077	08:22
Air Blank	0.000	08:23
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
05/27/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:24
Control Test	0.197	08:25
Air Blank	0.000	08:25
Control Test	0.198	08:26
Air Blank	0.000	08:26
Control Test	0.198	08:27
Air Blank	0.000	08:28
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
05/27/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:28
Control Test	0.079	08:29
Air Blank	0.000	08:29
Control Test	0.080	08:30
Air Blank	0.000	08:30
Control Test	0.080	08:30
Air Blank	0.000	08:31
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
BK Rel Std Dev(%)	0.7247	

DBS

Operator's Signature

FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Osceola Co SO S/N 80-000968

Date In 3/23/2015 Date Out 3/26/15 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED MAR 26 2015 FDLE Alcohol Testing Program

Intake Performed By SDN

- ☐ Registration
- ☑ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:

- ☑ Power cord
- ☐ Printer Cable
- ☐ Other

Notes:

Quality Checks Performed By ✓

- ☑ Breath Tube Screen
- ☑ Replace O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 195
- ☑ Flow Verification (L/s)

Flow Column # A78101

32mm 144 (.139 - .169) ✓
36mm 167 (.156 - .190) ✓
53mm 239 (.228 - .278) ✓
103mm 504 (.447 - .547) ✓

- ☑ Barometric Pressure Check

Gauge ID # 26932

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1024	2014 11 K / 11/25/16
0.08	DR2035	201502G / 2/24/17
0.20	DR3856	201462A / 2/11/16
0.08 DGS	N/A	AG429701 / 10/24/16

Flow Calibration Performed By

- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column #

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☑ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

- ☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By ✓

- ☑ Barometric Pressure ID# 26932 — 1014 Gauge, 1014 Instrument

Mouth Alcohol Solution Lot # 2014-A

Acetone Stock Solution Lot # 2014-B

Simulator	Serial Number
0.00	SD1016
Interferent	G-2840
0.05	SD1024
0.08	DR2035
0.20	DR3856

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: QC - BK

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

Patrick Murphy
Quality Control Review

3/26/15
Date

Revised November 2012

P8990

80-000968
Stability Checks
3/26/15

INTOXILYZER 8000
Instrument Initialization
08:24 03/26/2015

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
03/26/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:52
Control Test	0.079	08:52
Air Blank	0.000	08:53
Control Test	0.079	08:53
Air Blank	0.000	08:54
Control Test	0.079	08:54
Air Blank	0.000	08:54
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
03/26/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.198	08:56
Air Blank	0.000	08:57
Control Test	0.199	08:57
Air Blank	0.000	08:58
Control Test	0.197	08:59
Air Blank	0.000	08:59
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
03/26/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:01
Control Test	0.079	09:01
Air Blank	0.000	09:02
Control Test	0.079	09:02
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:04
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

OSCEOLA CORRECTIONS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000968
03/26/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:05
Control Test	0.048	09:06
Air Blank	0.000	09:07
Control Test	0.049	09:07
Air Blank	0.000	09:08
Control Test	0.049	09:08
Air Blank	0.000	09:09
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature