FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Satellite Beach PD S/N 80-001001

Date In 11/5/15 Date Out 11/10/15 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED
NOV 10 2015
FDLE
Alcohol Testing Program

Intake Performed By SW

- ☐ Registration
- ☒ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:
- ☐ Power cord
- ☐ Printer Cable
- ☒ Other Static bag

Notes:

Quality Checks Performed By

- ☒ Breath Tube Screen
- ☒ Replace O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 2.01
- ☒ Flow Verification (L/s)

Flow Column # ATP105

32mm	149	(.139 - .169)
36mm	163	(.156 - .190)
53mm	251	(.228 - .278)
103mm	503	(.447 - .547)

- ☒ Barometric Pressure Check
  Gauge ID # 26932
- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1024	20141K 11/25/16
0.08	DR1279	20150ZG 2/24/17
0.20	DR3856	20140SB 08/12/16
0.08 DGS	N/A	AB511Z01 4/27/17

Flow Calibration Performed By

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete
  Flow Column #
  - ☐ 5L/min – 17mm
  - ☐ 15L/min – 53mm
  - ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)
  Flow Column #

32mm	(.139 - .169)
36mm	(.156 - .190)
53mm	(.228 - .278)
103mm	(.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

- ☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By

☒ Barometric Pressure ID# 26932

1017	Gauge
1018	Instrument

Mouth Alcohol Solution Lot # 2015-A
Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1016
Interferent	G2540
0.05	SD1024
0.08	DR1279
0.20	DR3856

Attachments

- ☒ Form 41
- ☒ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: QC-BK

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

Patrick Murphy
Quality Control Review

11/10/15
Date

Revised November 2012

80-001001
Stability checks
11/10/15
B

INTOXILYZER 8000
Instrument Initialization
11:18 11/10/2015

RBM

SATELLITE BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001001
11/10/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:47
Control Test	0.079	11:47
Air Blank	0.000	11:48
Control Test	0.079	11:48
Air Blank	0.000	11:49
Control Test	0.079	11:49
Air Blank	0.000	11:50
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SATELLITE BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001001
11/10/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:51
Control Test	0.052	11:52
Air Blank	0.000	11:53
Control Test	0.051	11:53
Air Blank	0.000	11:54
Control Test	0.052	11:54
Air Blank	0.000	11:55
Control Test Stats		
Average	0.0517	
Std Dev	0.0006	
Rel Std Dev(%)	1.1175	

Operator's Signature

SATELLITE BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001001
11/10/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:57
Control Test	0.081	11:57
Air Blank	0.000	11:58
Control Test	0.082	11:59
Air Blank	0.000	11:59
Control Test	0.081	12:00
Air Blank	0.000	12:00
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

Operator's Signature

SATELLITE BEACH P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001001
11/10/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:02
Control Test	0.203	12:03
Air Blank	0.000	12:03
Control Test	0.203	12:04
Air Blank	0.000	12:05
Control Test	0.203	12:05
Air Blank	0.000	12:06
Control Test Stats		
Average	0.2030	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BK

Operator's Signature