FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Indialantic PD S/N 80-001006

Date In 11/25/15 Date Out 11/30/15 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED NOV 30 2015 FDLE Alcohol Testing Program

Intake Performed By SW

- ☐ Registration
- ☒ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other _____

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:

- ☒ Power cord
- ☐ Printer Cable
- ☒ Other Static bag

Notes: _____

Quality Checks Performed By [illegible]

- ☒ Breath Tube Screen
- ☒ Replace O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 215
- ☒ Flow Verification (L/s)

Flow Column # ATP103

32mm 152 (.139 - .169)
36mm 177 (.156 - .190)
53mm 253 (.228 - .278)
103mm 503 (.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 28427

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502G 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG511704 4/27/17

Flow Calibration Performed By _____

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32mm _____ (.139 - .169)
36mm _____ (.156 - .190)
53mm _____ (.228 - .278)
103mm _____ (.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Suggested Service

Optical Bench Calibration Performed By _____

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [illegible]

☒ Barometric Pressure 1019 Gauge
ID# 28427 1022 Instrument

Mouth Alcohol Solution Lot # 2015-A
Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

- ☒ Form 41
- ☒ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other _____

Notes: QCPWS

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

Brett Kirkland
Quality Control Review

11/30/15
Date

Revised November 2012

80-001006
Stability Checks
11/30/15

INTOXILYZER 8000
Instrument Initialization
08:10 11/30/2015

INDIALANTIC P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001006
11/30/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.050	09:10
Air Blank	0.000	09:11
Control Test	0.050	09:11
Air Blank	0.000	09:12
Control Test	0.051	09:13
Air Blank	0.000	09:13
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Operator's Signature

INDIALANTIC P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001006
11/30/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:15
Control Test	0.204	09:16
Air Blank	0.000	09:16
Control Test	0.205	09:17
Air Blank	0.000	09:18
Control Test	0.206	09:18
Air Blank	0.000	09:19
Control Test Stats		
Average	0.2050	
Std Dev	0.0010	
Rel Std Dev(%)	0.4878	

Operator's Signature

INDIALANTIC P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001006
11/30/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:20
Control Test	0.081	09:21
Air Blank	0.000	09:21
Control Test	0.080	09:22
Air Blank	0.000	09:23
Control Test	0.080	09:23
Air Blank	0.000	09:24
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

INDIALANTIC P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001006
11/30/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:25
Control Test	0.079	09:25
Air Blank	0.000	09:26
Control Test	0.079	09:26
Air Blank	0.000	09:26
Control Test	0.080	09:27
Air Blank	0.000	09:27
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DWS
BK
DGS

Operator's Signature