FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Haines City PD S/N 80-001007

Date In 6/9/15 Date Out 6/15/15 ☑ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By [illegible]

- ☐ Registration
- ☒ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement
- ☐ Other

Visual Inspection:

- Ok Case
- Ok Handle
- Ok Dry Gas Holder
- Ok Feet
- Ok Keyboard/Plug
- Ok Back/Plugs
- Ok Screws tight
- Ok Breath Hose

Other Equipment:

- ☐ Power cord
- ☐ Printer Cable
- ☐ Other

Notes:

Quality Checks Performed By ASM

- ☑ Breath Tube Screen
- ☑ Replace O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 193
- ☑ Flow Verification (L/s)
  - Flow Column # ATP.103
  - 32mm D.152 (.139 - .169)
  - 36mm D.1667 (.156 - .190)
  - 53mm D.234 (.228 - .278)
  - 103mm D.492 (.447 - .547)
- ☑ Barometric Pressure Check
  - Gauge ID # 28427
- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	ZD1411K 11/25/16
0.08	SD1011	201502G 2/24/17
0.20	SD1025	201408B 8/12/16
0.08 DGS	N/A	AG5026D3 1/26/17

Flow Calibration Performed By

- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete
  - Flow Column #
  - ☐ 5L/min – 17mm
  - ☐ 15L/min – 53mm
  - ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification(s)
  - Flow Column #
  - 32mm (.139 - .169)
  - 36mm (.156 - .190)
  - 53mm (.228 - .278)
  - 103mm (.447 - .547)

RECEIVED JUN 15 2015 FDLE Alcohol Testing Program

Maintenance Performed By ASM

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☑ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☑ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By ASM

- ☑ Barometric Pressure 1018 Gauge
  - ID# 28427 1020 Instrument

Mouth Alcohol Solution Lot # 2014-A

Acetone Stock Solution Lot # 2014-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☑ Other 2nd Form 41

Notes: QC - SDM

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

Patricia Murphy
Quality Control Review

6/15/15
Date

Revised November 2012

Stability Checks 80-001007 Haines City P.D. 6/15/15

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001007
06/15/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:59
Control Test	0.049	08:00
Air Blank	0.000	08:01
Control Test	0.050	08:01
Air Blank	0.000	08:02
Control Test	0.049	08:03
Air Blank	0.000	08:03
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

QC 005

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001007
06/15/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:04
Control Test	0.078	08:04
Air Blank	0.000	08:05
Control Test	0.079	08:06
Air Blank	0.000	08:06
Control Test	0.079	08:07
Air Blank	0.000	08:07
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001007
06/15/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:08
Control Test	0.199	08:09
Air Blank	0.000	08:09
Control Test	0.198	08:10
Air Blank	0.000	08:11
Control Test	0.198	08:11
Air Blank	0.000	08:12
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

Operator's Signature

DSS

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001007
06/15/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:13
Control Test	0.077	08:13
Air Blank	0.000	08:14
Control Test	0.077	08:14
Air Blank	0.000	08:15
Control Test	0.078	08:15
Air Blank	0.000	08:16
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

Operator's Signature