FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO S/N 80-001045

Date In 10/2/15 Date Out 10/13/15 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED OCT 16 2015 FDLE Alcohol Testing Program

Intake — Performed By TP

- ☐ Registration
- ☑ Annual
- ☑ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:
- ☑ Power cord
- ☐ Printer Cable
- ☑ Other Static Bag

Notes:

Quality Checks — Performed By BK

- ☒ Breath Tube Screen
- ☒ Replace O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 204
- ☒ Flow Verification (L/s)
  - Flow Column # ATP103
  - 32mm .152 (.139 - .169)
  - 36mm .167 (.156 - .190)
  - 53mm .238 (.228 - .278)
  - 103mm .496 (.447 - .547)
- ☒ Barometric Pressure Check
  - Gauge ID # 28427
- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	2014111K
		11/25/16
0.08	SD1011	2015020
		2/24/17
0.20	SD1025	2014088
		8/12/16
0.08 DGS	N/A	AG511701
		4/27/17

Flow Calibration — Performed By

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete
  - Flow Column # OCT
  - ☐ 5L/min – 17mm
  - ☐ 15L/min – 53mm
  - ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)
  - Flow Column #
  - 32mm (.139 - .169)
  - 36mm (.156 - .190)
  - 53mm (.228 - .278)
  - 103mm (.447 - .547)

Maintenance — Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration — Performed By

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ____ ID # ____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection — Performed By BK

☒ Barometric Pressure ID# 28427 — 1012 Gauge, 1011 Instrument

Mouth Alcohol Solution Lot # 2014-A

Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

- ☒ Form 41
- ☒ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: QC-

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

Patrick Murphy
Quality Control Review

10/16/15
Date

Revised November 2012

Jacksonville SO
80-001045
Stability Check
10/13/15 BK

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001045
10/13/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.050	09:48
Air Blank	0.000	09:49
Control Test	0.049	09:49
Air Blank	0.000	09:50
Control Test	0.050	09:51
Air Blank	0.000	09:51
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

BK
Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001045
10/13/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.079	09:59
Air Blank	0.000	09:59
Control Test	0.079	10:00
Air Blank	0.000	10:00
Control Test	0.079	10:01
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BK
Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001045
10/13/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.200	09:53
Air Blank	0.000	09:54
Control Test	0.202	09:55
Air Blank	0.000	09:55
Control Test	0.203	09:56
Air Blank	0.000	09:56
Control Test Stats		
Average	0.2017	
Std Dev	0.0015	
Rel Std Dev(%)	0.7574	

BK
Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001045
10/13/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.079	10:03
Air Blank	0.000	10:04
Control Test	0.078	10:04
Air Blank	0.000	10:05
Control Test	0.078	10:05
Air Blank	0.000	10:06
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DCS

BK
Operator's Signature