FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pasco Co SO S/N 80-001117

Date In 7/15/15 Date Out 7/20/15 ☐Ship ☑P/U ☐H/D ☐CMI ☐EE

RECEIVED JUL 21 2015 FDLE Alcohol Testing Program

Intake Performed By [illegible]

- ☐ Registration
- ☒ Annual
- ☒ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other _____

Visual Inspection:

- OK Case
- OK Dry Gas Holder
- OK Keyboard/Plug
- OK Screws tight
- OK Handle
- OK Feet
- OK Back/Plugs
- OK Breath Hose

Other Equipment:

- ☐ Power cord
- ☐ Printer Cable
- ☒ Other Static Bag

Notes: _____

Quality Checks Performed By [illegible]

- ☑ Breath Tube Screen
- ☑ Replace O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 135
- ☑ Flow Verification (L/s)
  - Flow Column # ATP105
  - 32mm 0.148 (.139 - .169)
  - 36mm 0.164 (.156 - .190)
  - 53mm 0.238 (.228 - .278)
  - 103mm 0.507 (.447 - .547)
- ☑ Barometric Pressure Check
  - Gauge ID # 28427
- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1098	201411K 11/25/16
0.08	SD1025	20150226 2/24/17
0.20	SD1011	201408B 8/12/16
0.08 DGS	N/A	AG50260D3 1/26/17

Flow Calibration Performed By [illegible]

- ☑ Flow Calibration N/A
- ☐ Flow Calibration Complete
  - Flow Column # _____
  - ☐ 5L/min – 17mm
  - ☐ 15L/min – 53mm
  - ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)
  - Flow Column # _____
  - 32mm _____ (.139 - .169)
  - 36mm _____ (.156 - .190)
  - 53mm _____ (.228 - .278)
  - 103mm _____ (.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Suggested Service

Optical Bench Calibration Performed By _____

- ☑ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [illegible]

- ☑ Barometric Pressure 1014 Gauge
  - ID# 28427 1012 Instrument

Mouth Alcohol Solution Lot # 2015-A

Acetone Stock Solution Lot # 2014-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1025
0.20	SD1011

Attachments

- ☑ Form 41
- ☑ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other _____

Notes: Oc-TSK

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

Patricia C Murphy
Quality Control Review

7/21/15
Date

Revised November 2012

Stability Checks 80-001117 Pasco County S.O. 7/20/15

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
07/20/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.053	09:39
Air Blank	0.000	09:40
Control Test	0.052	09:40
Air Blank	0.000	09:41
Control Test	0.051	09:42
Air Blank	0.000	09:42
Control Test Stats		
Average	0.0520	
Std Dev	0.0010	
Rel Std Dev(%)	1.9231	

Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
07/20/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.051	09:44
Air Blank	0.000	09:44
Control Test	0.051	09:45
Air Blank	0.000	09:45
Control Test	0.051	09:46
Air Blank	0.000	09:47
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
07/20/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.082	09:48
Air Blank	0.000	09:49
Control Test	0.082	09:50
Air Blank	0.000	09:50
Control Test	0.082	09:51
Air Blank	0.000	09:52
Control Test Stats		
Average	0.0820	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
07/20/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.205	09:53
Air Blank	0.000	09:54
Control Test	0.204	09:54
Air Blank	0.000	09:55
Control Test	0.205	09:56
Air Blank	0.000	09:56
Control Test Stats		
Average	0.2047	
Std Dev	0.0006	
Rel Std Dev(%)	0.2821	

BK

Operator's Signature

DGS

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
07/20/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:00
Control Test	0.080	10:01
Air Blank	0.000	10:01
Control Test	0.080	10:01
Air Blank	0.000	10:02
Control Test	0.081	10:02
Air Blank	0.000	10:03
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

Sims temporarily lost power. Allowed to sit, retested.