FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Seminole County PD S/N 80-001272

Date In 11/10/15 Date Out 11/18/15 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED NOV 18 2015 FDLE Alcohol Testing Program

Intake — Performed By SLS

- ☐ Registration
- ☒ Annual
- ☐ Return from CMI
- ☐ Return from Enforcement
- Electronics
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:
- ☐ Power cord
- ☐ Printer Cable
- ☒ Other Static bag

Notes:

Quality Checks — Performed By

- ☒ Breath Tube Screen
- ☒ Replace O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 131
- ☒ Flow Verification (L/s)

Flow Column # ATP102

32mm	144	(.139 - .169)
36mm	163	(.156 - .190)
53mm	240	(.228 - .278)
103mm	496	(.447 - .547)

- ☒ Barometric Pressure Check
  Gauge ID # 28427
- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	2015070 / 7/14/17
0.08	SD1011	2015026 / 2/24/17
0.20	SD1025	2015053A / 5/12/17
0.08 DGS	N/A	AG511701 / 4/27/17

Flow Calibration — Performed By

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column #
- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32mm	(.139 - .169)
36mm	(.156 - .190)
53mm	(.228 - .278)
103mm	(.447 - .547)

Maintenance — Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration — Performed By

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection — Performed By

- ☒ Barometric Pressure ID# 28427
  1015 Gauge
  1013 Instrument

Mouth Alcohol Solution Lot # 2015-A

Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

- ☒ Form 41
- ☒ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: QC-BK

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

Patrick Murphy
Quality Control Review

11/18/15
Date

Revised November 2012

80-001272
Stability Checks
11/18/15

INTOXILYZER 8000
Instrument Initialization
08:53 11/18/2015

SEMINOLE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001272
11/18/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.078	09:53
Air Blank	0.000	09:54
Control Test	0.079	09:54
Air Blank	0.000	09:54
Control Test	0.079	09:55
Air Blank	0.000	09:55
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

SEMINOLE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001272
11/18/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.204	09:59
Air Blank	0.000	10:00
Control Test	0.204	10:00
Air Blank	0.000	10:01
Control Test	0.203	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.2037	
Std Dev	0.0006	
Rel Std Dev(%)	0.2835	

Operator's Signature

SEMINOLE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001272
11/18/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.081	10:04
Air Blank	0.000	10:05
Control Test	0.081	10:05
Air Blank	0.000	10:06
Control Test	0.081	10:07
Air Blank	0.000	10:07
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SEMINOLE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001272
11/18/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.051	10:09
Air Blank	0.000	10:09
Control Test	0.050	10:10
Air Blank	0.000	10:10
Control Test	0.052	10:11
Air Blank	0.000	10:12
Control Test Stats		
Average	0.0510	
Std Dev	0.0010	
Rel Std Dev(%)	1.9608	

Operator's Signature