FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO S/N 80-001743

Date In 10/5/15 Date Out 10/23/15 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

RECEIVED OCT 26 2015 FDLE Alcohol Testing Program

Intake Performed By TP

- ☐ Registration
- ☑ Annual
- ☑ Return from CMI
- ☐ Return from Enforcement Electronics
- ☐ Other

Visual Inspection:

OK Case	OK Handle
OK Dry Gas Holder	OK Feet
OK Keyboard/Plug	OK Back/Plugs
OK Screws tight	OK Breath Hose

Other Equipment:

- ☑ Power cord
- ☐ Printer Cable
- ☑ Other Static Bag

Notes:

Quality Checks Performed By BK

- ☒ Breath Tube Screen
- ☒ Replace O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 238
- ☒ Flow Verification (L/s)

Flow Column # ATP103

32mm .148 (.139 - .169)
36mm .167 (.156 - .190)
53mm .238 (.228 - .278)
103mm .496 (.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 28427

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	2014111K 11/25/16
0.08	SD1011	2015026 2/24/17
0.20	SD1025	2014081B 8/12/16
0.08 DGS	N/A	A651701 4/27/17

Flow Calibration Performed By

- ☒ Flow Calibration N/A
- ☐ Flow Calibration Complete

Flow Column #

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value
- ☐ Post Calibration Verification (L/s)

Flow Column #

32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)

Maintenance Performed By

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

- ☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By BK

- ☒ Barometric Pressure 1018 Gauge
ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2014-A
Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD 1022
Interferent	SD 1021
0.05	SD 1018
0.08	SD 1011
0.20	SD 1025

Attachments

- ☒ Form 41
- ☒ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☐ Other

Notes: QC-PWS

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

Patrick Murphy
Quality Control Review

10/26/15
Date

Revised November 2012

PDM

Jacksonville SO
80-001743
stability Check
10/23/15 BK

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001743
10/23/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:29
Control Test	0.049	11:29
Air Blank	0.000	11:30
Control Test	0.050	11:31
Air Blank	0.000	11:31
Control Test	0.050	11:32
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

BK
Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001743
10/23/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:21
Control Test	0.080	11:22
Air Blank	0.000	11:23
Control Test	0.080	11:23
Air Blank	0.000	11:24
Control Test	0.080	11:24
Air Blank	0.000	11:25
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BK
Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001743
10/23/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.206	11:35
Air Blank	0.000	11:36
Control Test	0.205	11:36
Air Blank	0.000	11:37
Control Test	0.205	11:38
Air Blank	0.000	11:38
Control Test Stats		
Average	0.2053	
Std Dev	0.0006	
Rel Std Dev(%)	0.2812	

BK
Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001743
10/23/2015
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:47
Control Test	0.079	11:47
Air Blank	0.000	11:48
Control Test	0.080	11:48
Air Blank	0.000	11:48
Control Test	0.080	11:49
Air Blank	0.000	11:49
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

BK
Operator's Signature