

21:41:20 11-22-2016  
 11/23/16

RECEIVED  
 NOV 23 2016  
 FDLE  
 Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ALACHUA COUNTY SO  
 Time of Inspection: 08:41

Date of Inspection: 11/23/2016

Serial Number: 80-000760  
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5


Remarks:  
 Non-compliance: DID NOT COMPLETE TEST.

*I completed the alcohol free/mouth alcohol test and reviewed the results. when I hit enter to move to the alcohol free test it took me to review remarks. I said yes then hit enter and it took me to in compliance page I hit No. I then completed another Agency Inspection with no problems.*

*I contacted Patrick Murphy at 9:44am to let him know what happened.*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 TONI L FULTON  
 Signature and Printed Name

11/23/2016  
 Date