

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

DEC 27 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE CO SO
Time of Inspection: 08:02

Date of Inspection: 12/21/2016

Serial Number: 80-000768
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____


Remarks:
BYPASS AI TO OPERATE INSTRUMENT

RLS

**Compliance not determined*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAKE L SHANAHAN

Signature and Printed Name

12/21/2016
Date