


Florida Department of Law Enforcement Alcohol Testing Program


RECEIVED
 NOV 25 2016
 FDLE
 Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
 Time of Inspection: 18:18

Date of Inspection: 10/31/2016

Serial Number: 80-000770
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/06/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 14816080A3 Exp: 07/05/2018
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.199	0.079

Number of Simulators Used: 4

Remarks: *ON 0.05 SOLUTION EXP. DATE IS 3/8/2018 NOT 3/6/2018
(AMENDED)*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

W.C. Curry

 Signature and Printed Name WILLIAM C CURRY

10/31/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

SEP 08 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: HAMILTON COUNTY SO
Time of Inspection: 02:07

Date of Inspection: 08/29/2016

Serial Number: 80-000770
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06415080A2 Exp: 04/05/2017
0.000	0.049	0.074 / 0.079	0.199	0.077
0.000	0.049	0.075 / 0.080	0.200	0.077
0.000	0.050	0.079 / 0.080	0.199	0.077

Number of Simulators Used: 4

Remarks:
SIM HADNT WARMED TO PROPER TEMP,SENDING FOR ANNUAL INSPEC

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

William C Curry WILLIAM C CURRY
Signature and Printed Name

08/29/2016
Date