

RECEIVED

JUN 08 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLUMBIA COUNTY SO
Time of Inspection: 18:18

Date of Inspection: 05/24/2016

Serial Number: 80-000775
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1686772 Exp: 06/02/2017
0.000	0.045	0.076	0.201	0.080
0.000	0.046	0.077	0.200	0.081
0.000	0.045	0.077	0.201	0.081

Number of Simulators Used: 5


Remarks:

A F / M A: Ambient Fail.

too much mouth Alcohol Added for. solved issue

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
 Signature and Printed Name MICHAEL B CREWS

05/24/2016
Date

RECEIVED

Florida Department of Law Enforcement JUN 08 2016
Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLUMBIA COUNTY SO
Time of Inspection: 17:38

Date of Inspection: 05/24/2016

Serial Number: 80-000775
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000 / 0.000				

Number of Simulators Used: 5


Remarks:

00: Ambient Fail, Ambient Fail. Non-compliance:

AdS-A Fan - too much mouth alcohol

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL B CREWS
Signature and Printed Name

05/24/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 27 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: COLUMBIA COUNTY SO
Time of Inspection: 16:35

Date of Inspection: 01/24/2016

Serial Number: 80-000775
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1686772 Exp: 06/02/2017
/ 0.000	0.045	0.075	0.193	0.081
/ 0.000	0.046	0.076	0.196	0.082
/ 0.000	0.046	0.076	0.198	0.081

Number of Simulators Used: 5

Remarks:

00: Ambient Fail.

I moved the instrument out of the cabinet a little to fix the ambient fail.

Christopher Parker 1/24/16

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Christopher Parker 1/24/16

CHRISTOPHER A PARKER
Signature and Printed Name

01/24/2016
Date