

Florida Department of Law Enforcement Alcohol Testing Program

APR 18 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAFAYETTE CO SO
Time of Inspection: 10:42

Date of Inspection: 04/15/2016

Serial Number: 80-000776
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

SKIPPED AI TO OPERATE INSTRUMENT, NOT A COMPLIANCE CHECK

223

N/A Not a compliance check - PJM

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

04/15/2016
Date

RECEIVED

Florida Department of Law Enforcement JAN 14 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAFAYETTE CO SO
Time of Inspection: 10:37

Date of Inspection: 01/12/2016

Serial Number: 80-000776
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22315080A2 Exp: 09/05/2017
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.201	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Matthew Mandola
Signature and Printed Name MATTHEW MANDOLA

01/12/2016
Date