

RECEIVED

MAR 02 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 19:44

Date of Inspection: 01/29/2016

Serial Number: 80-000786
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1789170 Exp: 01/14/2018
0.000	0.050	0.078	0.200	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.050	0.079	0.201	0.079


Number of Simulators Used: 5

Remarks:

F. DICENZO WAS PRESENT & OBSERVED ENTIRE INSPECTION

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name TONY SUN

01/29/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JAN 04 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 20:59

Date of Inspection: 01/01/2016

Serial Number: 80-000786
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:

I was called out to B.U.I. Checkpoint due to Intox. 80-000786 not able to come out of Disabled mode due to A.I. not uploaded on Dec. 11th 2015. I had to take instrument out of disabled mode by doing an Agency inspection only as an entry to take instrument out of Disabled Mode. A. Frazier Completed Upload When Instrument was transported back to BAT Unit. FDLE was contacted and advised Jan. 4th, 2016.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Franca R. Dicienzo

FRANCA R DICIENZO

Signature and Printed Name

01/01/2016
Date