

Florida Department of Law Enforcement NOV 01 2016
 Alcohol Testing Program

FDLE
 Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
 Time of Inspection: 14:50

Date of Inspection: 10/25/2016

Serial Number: 80-000795
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.201	0.078
0.000	0.050	0.081	0.201	0.079

Number of Simulators Used: 5

Remarks:

ON LINE / MONTHLY

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome

JAMES A NEWSOME
 Signature and Printed Name

10/25/2016
 Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

SEP 08 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 15:02

Date of Inspection: 08/31/2016

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.051	0.081	0.199	0.079
0.000	0.051	0.081	0.199	0.078
0.000	0.051	0.081	0.199	0.078

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome

JAMES A NEWSOME
Signature and Printed Name

08/31/2016
Date

RECEIVED

Florida Department of Law Enforcement AUG 04 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 14:42

Date of Inspection: 07/28/2016

Serial Number: 80-000795
Software: 8100.27

Table with 3 columns: Check or Test, YES, NO. Rows include: Date and/or Time Adjusted, Diagnostic Check (Pre-Inspection): OK, Alcohol Free Subject Test: 0.000, Mouth Alcohol Test: Slope Not Met, Interferent Detect Test: Interferent Detect, Diagnostic Check (Post-Inspection): OK.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for three different samples.

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name: James A Newsome, Date: 07/28/2016

RECEIVED

Florida Department of Law Enforcement

Alcohol Testing Program

JUL 06 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 15:27

Date of Inspection: 06/29/2016

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 19615080A2 Exp: 09/05/2017
0.000	0.051	0.081	0.200	0.079
0.000	0.051	0.082	0.201	0.079
0.000	0.051	0.082	0.202	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome

JAMES A NEWSOME

Signature and Printed Name

06/29/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 07 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO

Time of Inspection: 14:53

Date of Inspection: 05/31/2016

Serial Number: 80-000795

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 19615080A2 Exp: 09/05/2017
0.000	0.051	0.081	0.200	0.078
0.000	0.051	0.082	0.201	0.078
0.000	0.051	0.082	0.202	0.078

Number of Simulators Used: 1

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome

JAMES A NEWSOME

Signature and Printed Name

05/31/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 06 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 15:17

Date of Inspection: 04/28/2016

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.047	0.083	0.201	0.078
0.000	0.047	0.083	0.202	0.078
0.000	0.047	0.083	0.202	0.078

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome

JAMES A NEWSOME

Signature and Printed Name

04/28/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 06 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: GULF COUNTY SO
Time of Inspection: 15:32

Date of Inspection: 03/31/2016

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19614080A2 Exp: 09/05/2017
0.000	0.049	0.079	0.197	0.000 / 0.078
0.000	0.051	0.080	0.198	0.078 / 0.079
0.000	0.050	0.080	0.198	0.078 / 0.079

Number of Simulators Used: 4

Remarks:

OP: DRY GAS HOSE NOT HOOKED UP.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome

JAMES A NEWSOME

Signature and Printed Name

03/31/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 03 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 13:57

Date of Inspection: 02/29/2016

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201403F Exp: 03/25/2016	0.08g/210L Test (g/210L) Lot#: 201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 19615080A2 Exp: 09/05/2017
0.000	0.048	0.078	0.193	0.078
0.000	0.049	0.079	0.194	0.078
0.000	0.049	0.079	0.195	0.078

Number of Simulators Used: 1

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome JAMES A NEWSOME
Signature and Printed Name

02/29/2016
Date

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

FEB 03 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXLYZER 8000

Agency: FLORIDA TROOP 10 Date of Inspection: 01/26/2016
Inspector: ALIS Station: 2113 Software: 2113.02

Agency Name	FLORIDA TROOP 10
Inspector Name	ALIS
Date of Inspection	01/26/2016
Station	2113
Software	2113.02
Device Model	INTOXLYZER 8000
Device Serial	1000
Device Location	Slope Not Met
Device Status	Interferent Detect
Device Calibration	OK

Device ID	0.05g/210L Test (g/210L) Lot#: 201402P Exp: 03/26/2016	0.03g/2.0L Test (g/210L) Lot#: 201404C Exp: 06/22/2016	0.20g/210L Test (g/210L) Lot#: 201403B Exp: 11/17/2016	0.04 g/g/210L Lot#: 201405A Exp: 11/03/2016
01111	0.025	0.073	0.116	0.121
01110	0.049	0.080	0.137	0.070
01112	0.018	0.050	0.197	0.070

Inspector Signature: _____
Agency Representative: _____

This device is a non-compliant (Y) does not comply () with the requirements of the Florida Department of Law Enforcement by the Agency of Law Enforcement in accordance with the provisions of Chapter 318.03, F.S.

James A. Newsome
Inspector

01/26/2016
ALIS