

Florida Department of Law Enforcement Alcohol Testing Program

**FDLE
Alcohol Testing Program**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DIXIE COUNTY SO
Time of Inspection: 09:19

Date of Inspection: 01/27/2016

Serial Number: 80-000810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 13115080A1 Exp: 06/05/2017
0.000	0.046	0.076	0.198	0.079
0.000	0.047	0.077	0.198	0.079
0.000	0.047	0.077	0.199	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MATTHEW MANDOLA

Signature and Printed Name

01/27/2016
Date