

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

DEC 21 2016

**FDLE**  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Serial Number: 80-000814

Time of Inspection: 07:04

Date of Inspection: 12/21/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.079	0.201	0.079
0.000	0.048	0.078	0.201	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Brad R Collito*

BRAD R COLLITO

Signature and Printed Name

12/21/2016

Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
NOV 23 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Serial Number: 80-000814

Time of Inspection: 06:14

Date of Inspection: 11/23/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRAD R COLLITO

Signature and Printed Name

11/23/2016  
Date

RECEIVED

JUN 21 2016

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Serial Number: 80-000814

Time of Inspection: 07:53

Date of Inspection: 06/17/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2016
0.000	0.048	0.078	0.196	0.081
0.000	0.048	0.078	0.197	0.081
0.000	0.048	0.079	0.197	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRAD R COLLITO

Signature and Printed Name

06/17/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

MAY 23 2016

**FDLE**  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 07:46

Date of Inspection: 05/20/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.048	0.077	0.196	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

#1148

TERESA STEVENS

Signature and Printed Name

05/20/2016  
Date

80-000914 *EMB*

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 12 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 09:39

Date of Inspection: 04/12/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014080A2 Exp: 07/05/2016
0.000	0.047	0.077	0.195	0.079
0.000	0.048	0.078	0.196	0.079
0.000	0.047	0.078	0.196	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Teresa Stevens*  
Signature

TERESA STEVENS

Signature and Printed Name

04/12/2016  
Date

RECEIVED

OCT 16 2016

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 08:15

Date of Inspection: 10/16/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	0.047	0.077	0.198	0.080
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt B. Collito*

BRAD R COLLITO

Signature and Printed Name

10/16/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

SEP 14 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 10:05

Date of Inspection: 09/14/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	0.048	0.077	0.197	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt B. Collito #1193*

BRAD R COLLITO

Signature and Printed Name

09/14/2016  
Date

RECEIVED

AUG 19 2016

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 05:27

Date of Inspection: 08/19/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	0.047	0.078	0.197	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.079	0.200	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Brad R Collito*

BRAD R COLLITO

Signature and Printed Name

08/19/2016  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
JUL 19 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 09:03

Date of Inspection: 07/19/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 12716080A7 Exp: 06/05/2018
0.000	0.047	0.078	0.195	0.081
0.000	0.048	0.077	0.196	0.081
0.000	0.048	0.078	0.197	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt B C 193*

BRAD R COLLITO

Signature and Printed Name

07/19/2016  
Date

RECEIVED

MAR 18 2016

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 11:21

Date of Inspection: 03/18/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.047	0.077	0.195	0.080
0.000	0.047	0.077	0.196	0.081
0.000	0.048	0.077	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TERESA STEVENS

Signature and Printed Name

03/18/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

FEB 20 2016

**FDLE**  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 08:33

Date of Inspection: 02/20/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.049	0.078	0.197	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.079	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TERESA STEVENS

Signature and Printed Name

02/20/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 22 2016

## AGENCY INSPECTION REPORT - INTOXILYZER 8100 <sup>FDLE</sup> Alcohol Testing Program

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 09:08

Date of Inspection: 01/22/2016

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201404C Exp: 04/08/2016	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:16014080A2 Exp: 07/05/2016
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.077	0.196	0.079
0.000	0.048	0.077	0.196	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TERESA STEVENS

Signature and Printed Name

01/22/2016  
Date