

Florida Department of Law Enforcement Alcohol Testing Program

MAR 28 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 13:06

Date of Inspection: 03/28/2016

Serial Number: 80-000815
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28814080A3 Exp: 11/05/2016
0.000	0.050	0.080	0.203	0.000 / 0.079
0.000	0.049	0.080	0.203	0.078 / 0.079
0.000	0.050	0.080	0.202	0.078 / 0.078

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

Amended 3/28/16
NOT connected *JW*

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES W JACOBSON

Signature and Printed Name

03/28/2016
Date