

80-000823 QWS

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
DEC 28 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 11:06

Date of Inspection: 12/28/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.049	0.078	0.195	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name MATTHEW C BOWDEN

12/28/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
NOV 30 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 09:15

Date of Inspection: 11/30/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.048	0.078	0.194	0.080
0.000	0.048	0.078	0.194	0.079
0.000	0.049	0.078	0.194	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt. J.S. DeSantis 819
J.S. DeSantis

MATTHEW C BOWDEN
Signature and Printed Name

11/30/2016
Date

RECEIVED

OCT 31 2016

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 11:52

Date of Inspection: 10/27/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.049	0.078	0.193	0.000 / 0.080
0.000	0.049	0.078	0.194	0.000 / 0.079
0.000	0.050	0.079	0.194	0.000 / 0.080

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance/HOSE DISCONNECTED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt. J. S. Desantis
J. Desantis 818

MATTHEW C BOWDEN
Signature and Printed Name

10/27/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

SEP 30 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 16:40

Date of Inspection: 09/27/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1766638 Exp: 11/05/2017
0.000	0.051	0.080	0.197	0.080
0.000	0.051	0.080	0.197	0.080
0.000	0.051	0.080	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BONDEN

SANDRA W CHESSHER

Signature and Printed Name

09/27/2016
Date

RECEIVED

SEP 02 2016

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 10:20

Date of Inspection: 08/29/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1766638 Exp: 11/05/2017
0.000	0.049	0.080	0.196	0.080
0.000	0.049	0.080	0.197	0.080
0.000	0.049	0.080	0.196	0.079

Number of Simulators Used: 5

Remarks:

A F / M A: TOO MUCH SOLUTION IN MOUTH. REPEATED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BOWDEN
M. Bowden 1100

SANDRA W CHESSHER *[Signature]*

Signature and Printed Name

08/29/2016

Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUL 27 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 10:08

Date of Inspection: 07/27/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201501F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1766638 Exp: 11/05/2017
0.000	0.048	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.079


Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BELLEFONTE


SANDRA W CHESSHER 
Signature and Printed Name

07/27/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 06 2016

Alcohol Testing Program
FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 13:26

Date of Inspection: 06/27/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:176663B Exp: 11/05/2017
0.000	0.047	0.078	0.194	0.078
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.079	0.196	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BARDEN


SANDRA W CHESSHER

Sgt. Chessher 950

Signature and Printed Name

06/27/2016
Date

80-000823 DB

RECEIVED JUN 02 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testin

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 15:45

Date of Inspection: 05/31/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1766638 Exp: 11/05/2017
0.000	0.047	0.079	0.193	0.079
0.000	0.048	0.079	0.194	0.079
0.000	0.049	0.079	0.195	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BOWDEN
[Signature]

Signature and Printed Name

SANDRA W CHESSHER

[Signature]

05/31/2016
Date

80-000823 WDB

RECEIVED

MAY 09 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 20:11

Date of Inspection: 04/30/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.047	0.079	0.196	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:
A F / M A: RFI Detect. REMOVED CELL PHONE FROM AREA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SANDRA W CHESSHER
Signature and Printed Name

04/30/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 09 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 15:06

Date of Inspection: 03/31/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.079	0.196	0.078
0.000	0.049	0.079	0.197	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

SANDRA W CHESSHER



03/31/2016
Date

RECEIVED

FEB 26 2016

FDLE
Alcohol Testing

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 10:17

Date of Inspection: 02/25/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.049	0.080	0.197	0.081
0.000	0.049	0.080	0.197	0.081
0.000	0.050	0.081	0.197	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BORDEN
[Signature]

SANDRA W CHESSHER *[Signature]*

Signature and Printed Name

02/25/2016
Date

80-000823 DM

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FEB 03 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 22:55

Date of Inspection: 01/28/2016

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG402702 Exp: 01/27/2016
0.000	0.048	0.080	0.197	0.079
0.000	0.048	0.080	0.197	0.079
0.000	0.049	0.080	0.198	0.079

Number of Simulators Used: 5

Remarks:

A F / M A: NOT ENOUGH SOLUTION IN MOUTH.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Bowden

SANDRA W CHESSHER

Signature and Printed Name

01/28/2016
Date