

827 RECEIVED

DEC 01 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 12:48

Date of Inspection: 11/22/2016

Serial Number: 80-000827
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
INT / INT				

Number of Simulators Used: 5

Remarks:

A F / M A: Interferent Detect Ambient Fail. 00: Interferent Detect, Interferent Detect. Non-compliance: .

FAILED INSPECTION, TAKEN OUT OF SERVICE

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JEFFERY A RAKER

Signature and Printed Name

11/22/2016
Date

MARION COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000827
 11/22/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.198	12:20
Air Blank	0.000	12:21
Control Test	0.197	12:21
Air Blank	0.000	12:22
Control Test	0.199	12:22
Air Blank	0.000	12:23
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

MARION COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000827
 11/22/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:26
Control Test	0.080	12:26
Air Blank	0.000	12:27
Control Test	0.080	12:28
Air Blank	0.000	12:28
Control Test	0.080	12:29
Air Blank	0.000	12:29
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MARION COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000827
 11/22/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:14
Control Test	0.050	12:15
Air Blank	0.000	12:15
Control Test	0.051	12:16
Air Blank	0.000	12:16
Control Test	0.052	12:17
Air Blank	0.000	12:18
Control Test Stats		
Average	0.0510	
Std Dev	0.0010	
Rel Std Dev(%)	1.9608	

827

Operator's Signature

Operator's Signature


Operator's Signature

827
②

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000827
11/22/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	AMB*	12:51
Air Blank	0.000	12:51

*Ambient Fail



Operator's Signature

827


MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000827
11/22/2016
Software: 8100.27

DRY CONTROL TEST

Test	g/210L	Time
-----	-----	-----
Air Blank	0.000	13:05
Control Test	0.081	13:05
Air Blank	0.000	13:06




Operator's Signature

827
①

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000827
11/22/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:10
Control Test	INT*	13:10
Air Blank	0.000	13:11
Control Test	INT*	13:12
Air Blank	AMB**	13:12
Air Blank	0.000	13:12

*Interferent Detect
**Ambient Fail



Operator's Signature

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JAN 06 2016

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 08:17

Date of Inspection: 01/06/2016

Serial Number: 80-000827
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
SKIPPED TO OPERATE INSTRUMENT

*N/A AI not conducted
compliance not determined* 22

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JENNIFER H JACOBS

Signature and Printed Name

01/06/2016
Date